



Sep 09, 2020

Thermal cycling of the PCR

philippe.bechtold¹¹ETHZ - ETH Zurich

In Development

dx.doi.org/10.17504/protocols.io.bk6jkzcn

philippe.bechtold

DOI

dx.doi.org/10.17504/protocols.io.bk6jkzcn

PROTOCOL CITATION

philippe.bechtold 2020. Thermal cycling of the PCR. **protocols.io**
<https://dx.doi.org/10.17504/protocols.io.bk6jkzcn>

LICENSE

This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](https://creativecommons.org/licenses/by/4.0/), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

CREATED

Sep 09, 2020

LAST MODIFIED

Sep 09, 2020

PROTOCOL INTEGER ID

41899

DISCLAIMER:

DISCLAIMER – FOR INFORMATIONAL PURPOSES ONLY; USE AT YOUR OWN RISK

The protocol content here is for informational purposes only and does not constitute legal, medical, clinical, or safety advice, or otherwise; content added to protocols.io is not peer reviewed and may not have undergone a formal approval of any kind. Information presented in this protocol should not substitute for independent professional judgment, advice, diagnosis, or treatment. Any action you take or refrain from taking using or relying upon the information presented here is strictly at your own risk. You agree that neither the Company nor any of the authors, contributors, administrators, or anyone else associated with protocols.io, can be held responsible for your use of the information contained in or linked to this protocol or any of our Sites/Apps and Services.

- 1 Incubation: ⌚ 00:10:00 at 55°C
- 2 Initial Denaturation ⌚ 00:03:00 at 92°C
- 3 45 cycles of the following:
⌚ 00:00:25 at 55°C

🕒 00:00:02 95°C

Fluorescence is measured at the end of each 55°C period.