



Aug 13, 2020

Universal sandwich enzyme linked immunosorbent assay for investigating protein-LA (SpLA) interactions with immunoglobulins using a SpG-HRP conjugate.

Angel A Justiz-Vaillant¹, Norma McFarlane-Anderson²¹University of the West Indies St. Augustine; ²University of West Indies. Mona Campus**1** Works for me dx.doi.org/10.17504/protocols.io.bjpzkmp6[University of the West Indies](#) angel.vaillant@sta.uwi.eduAngel Justiz-Vaillant
University of the West Indies St. Augustine

DOI

dx.doi.org/10.17504/protocols.io.bjpzkmp6

PROTOCOL CITATION

Angel A Justiz-Vaillant, Norma McFarlane-Anderson 2020. Universal sandwich enzyme linked immunosorbent assay for investigating protein-LA (SpLA) interactions with immunoglobulins using a SpG-HRP conjugate.. **protocols.io**
<https://dx.doi.org/10.17504/protocols.io.bjpzkmp6>

LICENSE

This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

CREATED

Aug 13, 2020

LAST MODIFIED

Aug 13, 2020

PROTOCOL INTEGER ID

40409

- 1 This ELISA is used to study the interaction of a protein-LA (SpLA) with different immunoglobulin preparations from mammalian species.
- 2 The 96 well microtitre plate is coated overnight at 4°C with 2 µg/µl per well of a mixture of protein-L and protein-A in carbonate-bicarbonate buffer pH 9.6.
- 3 Then plate is treated with bovine serum albumin solution and washed 4X with PBS-Tween.
- 4 50 µl of animal serum (1 mg/ml) is added and incubated for 1h at room temperature and the microplate is rewashed 4X with PBS-Tween.

- 5 Then 50 µl of peroxidase-labeled SpG conjugate diluted 1:5000 in PBS-non-fat milk is added to each well and incubated for 1h at RT. The plate is washed 4X with PBS-Tween.
- 6 50 µl of 4 mg/ml o-phenylenediamine solution (OPD) is added and the plate is incubated 15 minutes at RT in the dark.
- 7 The reaction is stopped with 50 µl of 3M H₂SO₄ solution.
- 8 The plate is visually assessed for the development of colour and read in a microplate reader at 492 nm.
- 9 A cut-off point should be calculated as the mean of the optical density of negative controls x 3. The higher the OD value the higher will be the affinity of SpLA to mammalian immunoglobulins.