•



Aug 21, 2020

© ELISA for quantification of IL-39 in human serum.

Angel A Justiz-Vaillant¹

¹University of the West Indies St. Augustine

In Development dx.doi.org/10.17504/protocols.io.bj4ekqte

University of the West Indies angel.vaillant@sta.uwi.edu

Angel Justiz-Vaillant University of the West Indies St. Augustine

ABSTRACT

Interleukins (IL) are a type of cytokine first thought to be expressed by leukocytes alone but have later been found to be produced by many other body cells. They play essential roles in the activation and differentiation of immune cells, as well as proliferation, maturation, migration, and adhesion. They also have pro-inflammatory and antiinflammatory properties. The primary function of interleukins is, therefore, to modulate growth, differentiation, and activation during inflammatory and immune responses. Interleukins consist of a large group of proteins that can elicit many reactions in cells and tissues by binding to high-affinity receptors in cell surfaces.

B lymphocytes mainly produce IL-39. It acts on neutrophils inducing their differentiation or expansion. [1]

Reference

1. Justiz Vaillant AA, Qurie A. Interleukin. In: Stat Pearls. Treasure Island (FL): Stat Pearls Publishing; June 12, 2019.

DOI

dx.doi.org/10.17504/protocols.io.bj4ekgte

PROTOCOL CITATION

Angel A Justiz-Vaillant 2020. ELISA for quantification of IL-39 in human serum.. protocols.io https://dx.doi.org/10.17504/protocols.io.bj4ekqte

LICENSE

□ This is an open access protocol distributed under the terms of the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

CREATED

Aug 21, 2020

LAST MODIFIED

Aug 21, 2020

PROTOCOL INTEGER ID

40806

- An anti-human IL-39 coating antibody is adsorbed onto the microwells by incubation overnight at 4°C with carbonatebicarbonate buffer.
- Add 50 µl of human serum. Human IL-39 present in the serum sample binds to antibodies adsorbed into the microwells.

mprotocols.io 08/21/2020

Citation: Angel A Justiz-Vaillant (08/21/2020). ELISA for quantification of IL-39 in human serum. https://dx.doi.org/10.17504/protocols.io.bj4ekqte

3	The microplate is blocked with 3% non-fat milk-PBS buffer and later wash to remove unbound proteins.
4	Fifty (50) µl of biotin-conjugated anti-human IL-39 antibody is added. The optimal dilution must be investigated.
5	The microplate is rewashed with PBS-Tween 20 buffer, pH 7.4.
6	One hundred µl of streptavidin-HRP conjugate is added and it binds to the biotin-conjugated anti-human IL-39 antibody. The optimal dilution of this conjugate must be investigated.
7	The plate is washed following incubation to remove the unbound Streptavidin-HRP.
8	Add 100 μl of 3',3',5',5'- tetramethylbenzidine (TMB; Sigma-Aldrich) into each well.
9	Incubate the microwells in the dark for 15 min.
9	Incubate the microwells in the dark for 15 min. A colored product is formed in proportion to the quantity of human IL-39 present in the sample or standard.
10	A colored product is formed in proportion to the quantity of human IL-39 present in the sample or standard.
10	A colored product is formed in proportion to the quantity of human IL-39 present in the sample or standard. The reaction is terminated by addition of 100 μ l 3M H2SO4 and the absorbance is measured at 450 nm. A standard curve is made from 7 human IL-39 standard dilutions and the human IL-39 sample concentration is