

May 22, 2024

Magnetic bead cleaning of PCR products with Cytiva Sera-Mag Select



Forked from [Magnetic bead cleaning of PCR products](#)

DOI

dx.doi.org/10.17504/protocols.io.n2bvj393wlk5/v1

Alexander Eiler¹, Even Garvang¹

¹University of Oslo

AQUA at UiO



Even Garvang

University of Oslo

OPEN  ACCESS



DOI: dx.doi.org/10.17504/protocols.io.n2bvj393wlk5/v1

Protocol Citation: Alexander Eiler, Even Garvang 2024. Magnetic bead cleaning of PCR products with Cytiva Sera-Mag Select. protocols.io <https://dx.doi.org/10.17504/protocols.io.n2bvj393wlk5/v1>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working

We use this protocol and it's working

Created: September 13, 2023

Last Modified: May 22, 2024

Protocol Integer ID: 87715



Disclaimer

DISCLAIMER – FOR INFORMATIONAL PURPOSES ONLY; USE AT YOUR OWN RISK

The protocol content here is for informational purposes only and does not constitute legal, medical, clinical, or safety advice, or otherwise; content added to **protocols.io** is not peer reviewed and may not have undergone a formal approval of any kind. Information presented in this protocol should not substitute for independent professional judgment, advice, diagnosis, or treatment. Any action you take or refrain from taking using or relying upon the information presented here is strictly at your own risk. You agree that neither the Company nor any of the authors, contributors, administrators, or anyone else associated with **protocols.io**, can be held responsible for your use of the information contained in or linked to this protocol or any of our Sites/Apps and Services.

Abstract

Magnetic bead cleaning of PCR products, adapted for Cytiva Sera-Mag Select beads. Forked from an older protocol using AMPure XP beads with minimal changes.

Guidelines

Work at post-PCR lab.

You can use PCR tubes/strips or PCR-plates but for easiest work with magnetic beads flat-bottom 96-well microplates are recommended.

- 1 Calculate the volume needed for the beads. For normal length primers a bead:sample ratio of 0.8:1 is likely OK. For long primers you can use a ratio of down to 0.65:1 (probably even lower). If you can still see primer dimers on a gel after cleanup, try a lower bead:sample ratio.

$0.8 \times \text{sample volume} =$


- 2 Shake the magnetic bead buffer bottle to fully resuspend magnetic particles.

- 3 Prepare 85% ethanol. You need 400 μL per sample, but prepare some extra.


Remember that when mixing ethanol and water, the total volume decreases (e.g. 15 ml water + 85 ml 100% ethanol < 100ml 85% ethanol). For the same reason, measure ethanol and water separately before mixing

- 4 Add Sample.

- 5 Add magnetic beads. Pipette mix 10 times and Incubate at room temperature for 5 minutes.

 00:05:00


- 6 Place the reaction plate onto a Magnet Plate for 2 minutes to separate beads from solution. (Keep on the magnet plate until step 11).

 00:02:00

- 7 Aspirate the supernatant from the reaction plate and discard.


- 8 Dispense 200 μL of 85% ethanol and incubate at room temperature for at least 30 seconds. Aspirate out the ethanol and discard. You can discard the ethanol by inverting the plate keeping the magnetic plate on the bottom of it.

 200 μL

 00:00:30

- 9 Repeat the washing step as above: dispense 200 μL of 85% ethanol and incubate at room temperature for at least 30 seconds. Aspirate out the ethanol and discard. You can discard the ethanol by inverting the plate keeping the magnetic plate on the bottom of it.

 200 μL

 00:00:30





10 Remove the remaining ethanol by pipetting.

11 Take the plate off the magnet



12 Dry off the remaining ethanol. To speed up the process, you can place the plate on a warm (30°C) heat block.

Drying takes at least 30 minutes. Check that drying was complete before continuing.

00:30:00

13 Add 40 μ L of elution buffer (PCR water or 10mM Tris), pipette mix 10 times. Make sure the buffer reaches up to the beads!

40 μ L

14 Incubate at room temperature for 2 minutes.

00:02:00

15 Place the reaction plate onto a Magnet Plate for 1 minute to separate beads from solution.



00:01:00

16 Transfer purified product to a new PCR plate or strips. Ensure that no beads are carried over when transferring, as it may inhibit later steps.