

Version 2 ▾

Aug 10, 2020

FFPE Tissue Pre-treatment Before t-CyCIF on Leica Bond RX V.2

Jia Ren Lin¹, Benjamin Izar^{1,2}, Zoltan Maliga^{1,2}, Yu-An Chen¹, Giorgio Gaglia^{1,2,3}, Ziming Du^{1,2}, Clarence Yapp¹, Shaolin Mei¹, Sandro Santagata^{1,2,3}, Peter Sorger^{1,2}

¹Laboratory of Systems Pharmacology, Harvard Medical School, Boston, MA;

²Ludwig Center for Cancer Research at Harvard, Harvard Medical School, Boston, MA;

³Department of Pathology, Brigham and Women's Hospital, Harvard Medical School, Boston, MA

1

Works for me

dx.doi.org/10.17504/protocols.io.bji2kkge

Laboratory of Systems Pharmacology

NCIHTAN



Madison Tyler

Laboratory of Systems Pharmacology

ABSTRACT

Tissue-based cyclic immunofluorescence (t-CyCIF) is optimized for FFPE specimens mounted on glass slides. Dewaxing and antigen retrieval are important steps to remove wax and expose antigenic sites. This protocol describes dewaxing and antigen retrieval on a Leica Bond RX automated slide processor; similar instruments are manufactured by Ventana or Dako and are commonly found in histopathology core facilities. t-CyCIF can also be performed following manual de-waxing and antigen retrieval (e.g. microwaving slides in citrate buffer or using a pressure cooker).

EXTERNAL LINK

www.cycif.org

DOI

dx.doi.org/10.17504/protocols.io.bji2kkge

PROTOCOL CITATION

Jia Ren Lin, Benjamin Izar, Zoltan Maliga, Yu-An Chen, Giorgio Gaglia, Ziming Du, Clarence Yapp, Shaolin Mei, Sandro Santagata, Peter Sorger 2020. FFPE Tissue Pre-treatment Before t-CyCIF on Leica Bond RX.

protocols.io<https://dx.doi.org/10.17504/protocols.io.bji2kkge>

MANUSCRIPT CITATION please remember to cite the following publication along with this protocol

EXTERNAL LINK

www.cycif.org

LICENSE

— This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](https://creativecommons.org/licenses/by/4.0/), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

CREATED

Aug 10, 2020

LAST MODIFIED

Aug 10, 2020

PROTOCOL INTEGER ID

40250

MATERIALS

NAME	CATALOG #	VENDOR
200 proof ethanol		
Epitope Retrieval Solution 1 BOND	AR9961	Leica Biosystems
Dewax Solution	AR9222	Leica Biosystems
Hoechst 33342	4082	Cell Signaling Technology
20X Phosphate Buffered Saline	28348	Thermo Fisher Scientific
Bond Universal Covertiles	S21.4611	Leica Biosystems
Intercept (PBS) Blocking Buffer	927-70001	LI-COR

MATERIALS TEXT


Fluorescently-conjugated secondary antibodies (experiment-specific)

EQUIPMENT

NAME	CATALOG #	VENDOR
Leica BOND RX	3342171	

1 Create a protocol file named "t-CyCIF" on the Leica Bond RX (see user manual for instrument-specific details):




- **CRITICAL STEP** We recommend creating a protocol file named "t-CyCIF" in the Leica Bond RX with the following conditions and saving for future use.






Leica BOND RX 3342171

<https://www.leicabiosystems.com/ihc-ish-fish/ihc-ish-instruments/products/leica-bond-rx/>

- 1.1 Bake FFPE slides at **60 °C** for **00:30:00**.
- 1.2 Dewax by rinsing three times with **150 µl** preheated (**60 °C**) Bond dewax solution, following by 3 rinses of **150 µl** , 200 proof ethanol.
- 1.3 Remove the Bond dewax solution. Add **150 µl** Bond ER1 solution for antigen retrieval and incubate at **99 °C** for **00:20:00**.
- 1.4 Remove the Bond ER1 solution and block for **00:30:00** with **150 µl** Intercept® (PBS) Blocking Buffer at **Room temperature**.


1.5 Remove the blocking buffer and incubate with  **150 µl** secondary antibody solution by incubating for  **01:00:00** at  **Room temperature** .


1.6 Remove the secondary antibody solution and incubate with  **150 µl** Hoechst solution for  **00:30:00** at  **Room temperature** .

2 Prepare reagent chambers for Leica Bond RX.

- **CRITICAL STEP** Prior to the first cycle of t-CyCIF, a prestaining (blocking) step is performed by incubating tissues with a mixture of appropriate secondary antibodies to block non-specific binding sites in the tissue. Secondary antibodies are chosen based on the species and isotypes of the unconjugated antibodies used in the first t-CyCIF cycle.
- **CRITICAL STEP** Avoid using Alexa Fluor 546-, Alexa Fluor 568- or Alexa Fluor 594-conjugated secondary antibodies, as these fluorophores are resistant to bleaching.

2.1 Fill chamber 1 with  **30 mL** of 1X PBS.

2.2 Fill chamber 2 with  **7 mL** of Intercept® (PBS) Blocking Buffer.

2.3 Fill chamber 3 with  **2 mL** of the appropriate secondary antibodies conjugated with Alexa Fluor 488, Alexa Fluor 555, or Alexa Fluor 647 diluted in Intercept® (PBS) Blocking Buffer (1:1000, vol:vol).

2.4 Open the lids of the chambers and put them in the chamber tray in the Leica Bond RX.

3 Place chambers in reagent tray and insert into Leica Bond RX.

4 Create a new study on the Leica Bond RX, select “t-CyCIF” file created as the protocol, add each slide to the study, and print labels for each FFPE slide. The Bond requires barcoded labels on all slides to process them.

- **CRITICAL STEP** Center the slide stickers evenly so that the Bond RX can scan the barcodes on the stickers correctly.

5 Place all labeled slides onto a slide tray, cover the slides with Bond Universal Covertiles, and insert the slide tray into the machine.

- **CRITICAL STEP** Be sure to place the Covertiles right-side-up, using the “Leica” etched on each Covertile as an orientation guide.

6 The machine will scan the barcodes on the stickers of the slides and reagent chambers. Check the label readings by

hovering the mouse over the slide labels. If all of the slides have been recognized correctly, then click the START button to run the protocol. This will take approximately 🕒 04:00:00 .

- 7 Remove slides from the Leica Bond RX and place them into 1X PBS.

■ **PAUSE POINT** Slides can be stored in 1X PBS at 🌡 4 °C for several days after processing on the Bond RX. Ensure that the entire tissue is covered in 1X PBS; otherwise the tissue will dry out and yield poor results.