

Jul 23, 2024

## BAF\_S07\_SpeedVac SPD1030

DOI

**[dx.doi.org/10.17504/protocols.io.j8nlk8mmwl5r/v1](https://dx.doi.org/10.17504/protocols.io.j8nlk8mmwl5r/v1)**

Nicholas Sherman<sup>1</sup>

<sup>1</sup>University of Virginia Biomolecular Analysis Facility Core

Biomolecular Analysis Fac...



Dilza Silva

UVA

---

OPEN  ACCESS



DOI: **[dx.doi.org/10.17504/protocols.io.j8nlk8mmwl5r/v1](https://dx.doi.org/10.17504/protocols.io.j8nlk8mmwl5r/v1)**

**Protocol Citation:** Nicholas Sherman 2024. BAF\_S07\_SpeedVac SPD1030. **protocols.io**

**<https://dx.doi.org/10.17504/protocols.io.j8nlk8mmwl5r/v1>**

**License:** This is an open access protocol distributed under the terms of the **[Creative Commons Attribution License](#)**, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

**Protocol status:** Working

**We use this protocol and it's working**

**Created:** July 13, 2024

**Last Modified:** July 23, 2024

**Protocol Integer ID:** 103373

## Abstract

General use of the SpeedVac for drying samples that contain aqueous or organic solvents. This is room temp and does not keep samples frozen like lyophilization.



## Materials

SpeedVac 1030 - Integrated Vacuum Concentrator - Thermo Scientific - part# SPD1030-115

Glass Condensation Flask- Thermo Scientific - part# GCF400

CryoCool™ (SCC1) Heat Transfer Fluid - Thermo Scientific - part# SCC1

Rotor: Savant 120 x 1.5 mL Vac Conc RTR - Thermo Scientific - part# RH12011

## SpeedVac SPD1030 ThermoScientific

### 1 **First start day:**

--> Ensure that the refrigerated trap contains a clean, dry, Glass Condensation Flask (GCF400) and that the supply of Thermo Scientific CryoCool™ (SCC1) Heat Transfer Fluid is sufficient. The CryoCool™ in the refrigerated trap must be cold before drying runs can begin. Turn unit on at least 45 minutes before starting the drying run.

For best results, in our Core we maintain electrical power to the system (the main power switch on the front of the unit in the lower right is "ON" and the front display is illuminated) at all the times to keep the refrigerated trap cold and ready for use.

## Operating

### 2 **Two modes of operation: with or without a fixed run time:**

**Manual run** - Starts a "Manual" run based on the parameters loaded in the screen, use manual run function if you do not want to set a fixed run time.

**Auto run** - Starts an "Automated" run based on the parameters loaded in the screen, use auto run function if you want to set a fixed run time.

### 3 **Auto run for peptides in 70% ACN/0.1% F.A or 0.1% TFA:**

The unit is turned on and the cover is unlocked.

The display is lighting up and shows the following in sequence:

Set values:

Temperature 45°C

Run time 2.00 h

Vacuum Pressure 5.1 Torr

Ramp vacuum: 5

Open cover, add samples with the lids open and organized samples to be equilibrated.

Close cover

Click Auto-run

Pay attention to listen cover lock sound and air removal.

Wait few minutes and make sure the vacuum pressure is going down.

### 4 **Auto run for metabolites in methanol/water:**

The unit is turned on and the cover is unlocked.

The display lights up and shows the following in sequence:

Set values:

Temperature 45°C

Run time 4.00 h

Vacuum Pressure 5.1 Torr

Ramp vacuum: 5



Open cover, add samples with the lids open and organized samples to be equilibrated.

Close cover

Click Auto-run

Pay attention to listen cover lock sound and air removal

Wait few minutes and make sure the vacuum pressure is going down.

**5 Auto-stop sound during auto run mode:**

1 beep that is repeated until the user pushes the **Stop** button or opens the cover of the concentrator

**6 Removing samples:**

Click stop button

Open cover

Remove samples

Close cover