




FEB 02, 2024

## F-2 FECES PROCESSING

REDI-NET Consortium<sup>1</sup><sup>1</sup>REDI-NET Consortium

REDI-NET Consortium

University of Notre Dame, Naval Medical Research Center, Wal...

### DISCLAIMER

This work is supported by the US Army Medical Research and Development Command under Contract No.W81XWH-21-C-0001, W81XWH-22-C-0093 and HT9425-23-C-0059. The views, opinions and/or findings contained in this report are those of the author(s) and should not be construed as an official Department of the Army or Navy position, policy or decision unless so designated by other documentation.

### ABSTRACT

This protocol details feces processing.

OPEN  ACCESS

DOI:

[dx.doi.org/10.17504/protocols.io.kxygx9ddog8j/v1](https://dx.doi.org/10.17504/protocols.io.kxygx9ddog8j/v1)

**Protocol Citation:** REDI-NET Consortium 2024. F-2 FECES PROCESSING. **protocols.io** <https://dx.doi.org/10.17504/protocols.io.kxygx9ddog8j/v1>

**License:** This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

**Protocol status:** Working  
We use this protocol and it's working

**Created:** Apr 28, 2023

Last Modified: Feb 02, 2024

PROTOCOL integer ID: 81132

**Keywords:** SAMPLE LYSIS, EXTRACTION, QUANTIFICATION AND STORAGE

**Funders Acknowledgement:**  
USAMRAA  
Grant ID: W81XWH-21-C-0001  
USAMRAA  
Grant ID: W81XWH-22-C-0093  
USAMRAA  
Grant ID: HT9425-23-C-0059

## GUIDELINES

### OBJECTIVE

To outline procedures for total nucleic acid extraction from fecal samples.

### SUMMARY/SCOPE

The overarching aim of the *REDI-NET* is to develop a collaborative laboratory network between domestic and international partnering institutions to address disease surveillance needs in order to effectively detect, predict and contain potentially emergent zoonosis. This SOP provides guidance on procedures for total nucleic acid extraction from fecal samples to provide materials for downstream library preparation and sequencing for pathogen detection.

### RESPONSIBLE PERSON

Principal Investigator, Study Coordinator, Entomology Component Lead, Managers

#### Note




**NOTE:** All study procedures must be conducted in compliance with national and local policies for prevention and control of COVID-19 infection.




### MAINTENANCE OF EQUIPMENT

#### Caution on RNA handling

1. RNases are very stable and difficult to inactivate, and only minute amounts are sufficient to destroy RNA.
2. Care should be taken to avoid inadvertently introducing RNases into the samples during or after the purification procedure.
3. Sample handling and extraction should be performed under an extraction hood and respecting Good Laboratory Practices.
4. Use filter tips all the time.


#### Storage of the buffers from IndiMag Pathogen Kit

1. Proteinase K is stable for at least 1 year after delivery when stored at  Room temperature (15-25°C). To store for more than 1 year or if ambient temperature often exceeds  25 °C, storage at  2-8 °C is recommended. Do not

- add Proteinase K directly to the Buffer VXL mixture! This can cause clogs or precipitates.
2. Precipitation may form after storage at low temperature or prolonged storage. To dissolve precipitate, incubate Buffer VXL or ACB for  00:30:00 at  37 °C, with occasional shaking.
3. Reconstituted Buffer AW1 can be stored at  Room temperature (15-25°C) for up to 1 year. Mix well after adding Ethanol.
4. Buffer AVE is RNase-free upon delivery. It contains sodium azide, an antimicrobial agent that prevents growth of RNase-producing organisms. However, as this buffer does not contain any RNase degrading chemicals, it will not actively inhibit RNases introduced by inappropriate handling. When handling Buffer AVE, take extreme care to avoid contamination with RNases. Follow general precautions for working with RNA, such as frequent change of gloves and keeping tubes closed whenever possible.

## APPENDICES

### APPENDIX 1. MEASURING SPOON FOR 0.1 MM BEATING BEADS

The spoon (Next Advance, MSP01-RNA) is used for 0.1 mm beating beads measurement. One scoop equals to  100 µL.



	Sample	Amount	Sample condition	Elution volume	DNA conc. (ng/ul)	RNA conc. (ng/ul)
	Fecal	100 mg	Frozen/Fresh	75	5-50	5 - 100

## MATERIALS

### EQUIPMENT AND MATERIALS

#### Note

**NOTE:** If product number is listed, please ensure use of this or equivalent product.

A	B
Equipment	Mfg / Product #
KingFisher™ Flex Magnetic Particle Processor with 96 Deep-Well Head or KingFisher™ Duo Prime Magnetic Particle Processor	ThermoFisher, 5400630 or ThermoFisher, 5400110
Bullet Blender 24 Gold	Next Advance, BB24-AU
Qubit 4 Fluorometer	ThermoFisher, Q33238
Adjustable micropipettes	Locally sourced
Multi-channel micropipettes	Locally sourced
Vortex	Locally sourced
Tube centrifuge	Locally sourced
Plate centrifuge	Locally sourced
Digital scale/balance	Locally sourced
Thermo Heater Mixer	Locally sourced

A	B	C
Material	Description	Mfg / Product #
IndiMag Pathogen Kit	w/o plastics, 384 reactions	Indical Bioscience, SP947257
Buffer ASL	140 mL, Stool Lysis Buffer	Qiagen, 19082
Measuring Spoon 100 µL	RNase Free, pack of 10. reusable	Next Advance, MSP01-RNA
KingFisher™ Deepwell 96 Plate	KingFisher	ThermoFisher, 95040450
KingFisher™ 96 KF microplate	KingFisher Flex <i>ONLY</i>	ThermoFisher, 97002540
KingFisher™ 96 tip comb for DW magnets	KingFisher Flex <i>ONLY</i>	ThermoFisher, 97002534
KingFisher™ Duo Prime 12-tip comb	KingFisher Duo Prime <i>ONLY</i>	ThermoFisher, 97003500


A	B	C
Elution Strip	KingFisher Duo Prime <i>ONLY</i>	ThermoFisher, 97003520
KingFisher™ Duo Cap for Elution Strip	KingFisher Duo Prime <i>ONLY</i>	ThermoFisher, 97003540
MicroAmp™ Clear Adhesive Film	KingFisher	ThermoFisher, 4306311
RNase-Free Microfuge Tubes	Nonstick, 1.5 mL	ThermoFisher, AM12450
Thermo Scientific Screw Cap Micro Tubes	1.5 mL Screw Cap Tube, NonKnurl, NonSkirted, Natural, E-Beam Sterile tube w/ attached cap	Fisher Scientific, 14-755-208
Zirconium oxide beads	0.1 mm, 400 g	Fisher Scientific, 50-154-2950
Qubit™ 1X dsDNA HS Assay Kit	(consumable)	ThermoFisher, Q33230
Qubit™ RNA HS Assay Kit	(consumable)	ThermoFisher, Q32852
Qubit Assay Tubes	For Qubit DNA/RNA measurement (consumable)	Thermo Fisher, Q32856
RNaseZap™ RNase Decontamination Solution	To remove RNase from working area	ThermoFisher, AM9780
ZymoBIOMICS Microbial Community Standard Material	For positive controls	Zymo Research, D6300
Zika virus (ZIKV) positive control	For TNA extraction positive control	NMRC made
Human gammaherpesvirus (EBV) positive control	For TNA extraction positive control	NMRC made
Forceps	For use with samples	Locally sourced
Spatula	For use with samples	Locally sourced
Ethanol	100% (molecular biology grade)	Locally sourced
Isopropanol	100% (molecular biology grade)	Locally sourced
Nuclease-free Water	To elute total nucleic acids	Locally sourced
Dry ice	To maintain cold chain during sample handling	Locally sourced
Ice bucket	To contain the dry ice	Locally sourced
Kimwipes	To dry material	Locally sourced
Falcon tubes	15 mL and 50 mL	Locally sourced
Data sheets	REDI-NET DCS SP-1 Sample Processing Form	REDI-NET Data Portal

Equipment	
KingFisher™ Flex Purification System, KingFisher with 96 Deep-well Head	NAME
Purification System	TYPE
Thermo Scientific™	BRAND
5400630	SKU
<a href="https://www.thermofisher.com/order/catalog/product/5400630">https://www.thermofisher.com/order/catalog/product/5400630</a>	LINK


Equipment	
KingFisher™ Duo Prime Purification System	NAME
Purification System	TYPE
Thermo Scientific™	BRAND
5400110	SKU
<a href="https://www.thermofisher.com/order/catalog/product/5400110?SID=srch-srp-5400110">https://www.thermofisher.com/order/catalog/product/5400110?SID=srch-srp-5400110</a>	LINK

Equipment	
Bullet Blender 24 Gold (1.5 mL snap and screw cap tubes, 4°C cooling)	NAME
Blender	TYPE
Next Advance	BRAND
BB24-AU	SKU
<a href="https://www.nextadvance.com/product/bullet-blender-24-gold/">https://www.nextadvance.com/product/bullet-blender-24-gold/</a>	LINK

Equipment	
Qubit Fluorometer	NAME
Fluorometer	TYPE
Invitrogen	BRAND
Q33238	SKU
<a href="https://www.thermofisher.com/order/catalog/product/Q33238#/Q33238">https://www.thermofisher.com/order/catalog/product/Q33238#/Q33238</a>	LINK

 IndiMag Pathogen Kit w/o plastics (384 reactions) **INDICAL BIOSCIENCE Catalog #SP947257**

 Buffer ASL (560 ml) **Qiagen Catalog #19082**

 Measuring Spoon 100 uL RNase Free pack of 10 **Next Advance Catalog #MSP01-RNA**



KingFisher<sup>®</sup>; Flex<sup>®</sup>; Systems Consumables, KingFisher Flex Microtiter Deepwell 96 plate, V-bottom **Thermo Fisher Catalog #95040450**

KingFisher Microplate **Thermo Fisher Scientific Catalog #97002540**

KingFisher<sup>®</sup>; Flex<sup>®</sup>; Systems Consumables, KingFisher 96 tip comb for DW magnets **Thermo Fisher Catalog #97002534**

KingFisher<sup>®</sup>; Duo and KingFisher<sup>®</sup>; Duo Prime Consumables, 12-tip comb, for Microtiter 96 Deepwell plate **Thermo Fisher Catalog #97003500**

KingFisher<sup>®</sup>; Duo and KingFisher<sup>®</sup>; Duo Prime Consumables, Elution strip **Thermo Fisher Catalog #97003520**

KingFisher<sup>®</sup>; Duo and KingFisher<sup>®</sup>; Duo Prime Consumables, KingFisher Duo Cap for elution strip **Thermo Fisher Catalog #97003540**

MicroAmp<sup>®</sup>; Clear Adhesive Film **Thermo Fisher Catalog #4306311**

Nonstick, RNase-free Microfuge Tubes, 1.5 mL **Thermo Fisher Catalog #AM12450**

Thermo Scientific<sup>™</sup> Screw Cap Micro Tubes **Fisher Scientific Catalog #14-755-208**

Bertin Corp 0.1mm Zirconium oxide beads (450g) (qty 500) **Fisher Scientific Catalog #50-154-2950**

Qubit<sup>™</sup> 1X dsDNA High Sensitivity (HS) and Broad Range (BR) Assay Kits **Thermo Fisher Scientific Catalog #Q33230**

Qubit RNA HS (High Sensitivity) assay **Thermo Fisher Scientific Catalog #Q32852**

Qubit<sup>™</sup> Assay Tubes **Invitrogen - Thermo Fisher Catalog #Q32856**

RNaseZap<sup>™</sup> RNase Decontamination Solution **Thermo Fisher Scientific Catalog #AM9780**

ZymoBIOMICS Microbial Community Standard **Zymo Research Catalog #D6300**

## SAFETY WARNINGS






### RISK AND PERSONAL PROTECTION

1. Caution should be taken while processing samples as some chemicals may be harmful. Please use a fume-hood when required to avoid inhaling harmful chemicals.
2. Gloves should be worn all the time when handling samples.
3. Decontaminants such as DNA/RNaseZap could irritate the skin, avoid contact with skin while preparing the workbench for nucleic acid extractions.


## BEFORE START INSTRUCTIONS

### BEFORE START

1. Pre-cool the Bullet Blender by adding dry ice into the cooling compartment and running the cooling program.
2. Pre-heat the heater mixer at  95 °C.
3. Clean the work surfaces with RNaseZap, then wipe the surfaces with 70% /molecular biology grade ethanol to remove additional contaminants.
4. Transfer 0.1 mm zirconium oxide beads (2 spoons, Appendix 1) to Thermo Scientific Screw Cap Micro 1.5 ml Tubes.
5. For the first time use of IndiMag pathogen kit, add 100% ethanol to Buffer AW1 and AW2, and add 100% isopropanol to ACB as indicated on the bottles.
6. MagAttract Suspension G from IndiMag pathogen kit needs to be vortexed thoroughly for  00:03:00 (before first use) or  00:01:00 (before subsequent uses) to ensure that the magnetic silica particles are fully resuspended.
7. Aliquot nuclease-free water in big bottle into a few 15 ml tubes for preparing TNA elution in KingFisher Flex or KingFisher Duo Prime to avoid cross-contamination.

## 1. SAMPLE LYSIS

1

Add  700 µL of ASL buffer to the bead tubes prepared in step 4 from Before Start under the Guidelines & Warning tab.



2 Use a clean spatula or forceps to weigh about  $\text{100 mg}$  fecal sample (roughly 5 drops of mouse feces) and place it into each prepared bead tube (If fecal sample is frozen, thaw and keep it  $\text{On ice}$ ). Record the weight.



3 Include a positive control for each batch of samples: transfer  $\text{37.5 }\mu\text{L}$  ZymoBIOMICS Microbial Community Standard Material,  $\text{100 }\mu\text{L}$  EBV, and  $\text{100 }\mu\text{L}$  ZIKV standard into a tube from step 4 from Before Start. Add  $\text{250 }\mu\text{L}$  ASL buffer.

4 Include a negative control for each batch of samples: a bead tube from step 4 from Before Start with  $\text{500 }\mu\text{L}$  ASL buffer only.

5 Refill the dry ice compartment of the Bullet Blender if necessary. Load all samples into the Bullet Blender.

6 Set the speed at 12 and the time at 3. Press Start.

7 After beating, incubate samples at  $\text{95 }^{\circ}\text{C}$  for  $\text{00:15:00}$ .

15m



8 After incubation, beat the samples in the Bullet Blender (Refill the dry ice compartment if necessary) at speed 12 and time at 3.

9 Let the samples settle for  $\text{00:01:00}$  in the Bullet Blender and then repeat step 8.

1m



2. INSTRUMENT SET UP

10

Note

KingFisher Flex only, if using KingFisher Duo Prime, go to 6. INSTRUMENT SET UP.

Confirm 96 deep-well magnetic heads and 96 well deep-well heat blocks are being used.

11

Ensure the program **IndiMag\_Pathogen\_KF\_Flex\_4wash** or the program has been downloaded and loaded onto the KingFisher Flex instrument.

3. SET UP THE PROCESSING PLATES

12


Set up the Tip Comb, Wash, and Elution Plates outside the instrument according to the following table.

Note


**NOTE:** DO NOT use the elution buffer provided by the kit for TNA elution. The ingredients in the elution buffer inhibit the downstream DNA sequencing efficiency.

A	B	C	D	E
Plate ID	Plate position	Plate type	Reagent	Volume per well
Tip comb	7	Place a 96 Deep-well Tip comb in a deep-well plate		
Elution	6	Deep-Well	Nuclease-free water	75 µL
Wash 4	5	Deep-Well	100% ethanol	750 µL
Wash 3	4	Deep-Well	80% ethanol	750 µL
Wash 2	3	Deep-Well	Buffer AW2	700 µL
Wash 1	2	Deep-Well	Buffer AW1	700 µL
Sample	1	Sample Lysate	Lysate and lysis buffer	985 µL


## 4. EXTRACTION







13 Centrifuge the bead tubes with lysate from step 9 for  12000 x g, 00:05:00 . 5m



14 Add  20  $\mu\text{L}$  of Proteinase K into wells of a Deep-well plate (based on the number of samples).



15 Transfer  270  $\mu\text{L}$  supernatant without any particle carryover to the wells of the Deep-well plate containing proteinase K. This plate becomes the Sample Plate.

16 Add  135  $\mu\text{L}$  Buffer VXL,  540  $\mu\text{L}$  Buffer ACB, and  20  $\mu\text{L}$  MagAttract Suspension G to each 2m  
sample in the sample plate. For multiple samples, make a master mix with 10% overage. Invert slowly to mix  
the master mix; avoid foaming (it can be mixed on a Hula mixer for  00:02:00 ). Add  695  $\mu\text{L}$   
mixture to each sample. Each well including sample lysate should be  985  $\mu\text{L}$  .



17 Select the program **IndiMag\_Pathogen\_KF\_Flex\_4wash** on the instrument.

18 Start the run, then load the prepared plates into position when prompted by the instrument.

## 5. QUANTIFICATION AND STORAGE

19 After the running protocol is completed (~35 minutes), immediately remove the elution plate from the instrument and cover the plate or transfer the eluate to the final tube or plate of choice for final storage.

- 20 In a 0.6-mL microcentrifuge tube, use  $1\ \mu\text{L}$  total nucleic acid for DNA and RNA concentration measurement using Qubit 4 Fluorometer following manufacturer instructions.

#### Note

**Kits needed:** Qubit 1X dsDNA HS Assay Kit and Qubit RNA HS Assay Kit. (see Appendix 2 and Appendix 3).

**NOTE:** If the DNA or RNA concentration measurement is too high, dilute  $2\ \mu\text{L}$  of sample in a new 1.5-mL tube with  $18\ \mu\text{L}$  nuclease-free water. Use  $3\ \mu\text{L}$  of the diluted sample for DNA or RNA quantification.

- 21 Proceed with sample testing following the REDI-NET SOP F-4 Fecal Testing or store at  $-20\ ^\circ\text{C}$  for less than 2 weeks.

#### Note

For long-term storage the sample needs to be stored at  $-80\ ^\circ\text{C}$  following the REDI-NET SOP F-3 Fecal Storage.

## 6. INSTRUMENT SET UP

22

#### Note

KingFisher Duo Prime only, if using KingFisher Flex, go to 2. INSTRUMENT SET UP.

Confirm 12-tip magnetic heads and 12 well deep-well heat blocks are being used.

- 23 Ensure the program IndiMag\_Pathogen\_KF\_Duo\_4wash has been downloaded and loaded onto the KingFisher Duo Prime instrument.

## 7. SET UP THE SAMPLE PLATE AND ELUTION STRIP

- 24 Set up the Sample Plate according to the table below:

A	B	C	D
Row ID	Plate Row	Reagent	Volume per well
Sample row	A	Lysate and lysis buffer	985 µL
Wash 1	B	Buffer AW1	700 µL
Wash 2	C	Buffer AW2	700 µL
Wash 3	D	80% ethanol	750 µL
Wash 4	E	100% ethanol	750 µL
	F	Tip comb	
	G	Empty	
	H		


25 Set up the Elution Strip according to the table below:

#### Note

**NOTE:** DO NOT use the elution buffer provided by the kit for TNA elution. The ingredients in the elution buffer inhibit the downstream DNA sequencing efficiency.

A	B	C	D
Strip ID	Row	Reagent	Volume per well
Elution	A	Nuclease-free water	75 µL

## 8. EXTRACTION


26 Centrifuge the bead tubes with lysate from step 9 for  12000 x g, 00:05:00 .






5m



27 Add  20 µL of Proteinase K into wells of Row A (based on number of samples) of a new Deep-well plate.



28 Transfer  270 µL supernatant without any particle carryover to the wells of the Deep-well Row A containing proteinase K. This row becomes the Sample Row.

29 Add  135 µL Buffer VXL,  540 µL Buffer ACB, and  20 µL MagAttract Suspension G to each sample in the sample plate. For multiple samples, make a master mix with 10% overage. Invert slowly to mix the master mix, avoid foaming. Add  695 µL mixture to each sample. Each well including sample lysate should be  985 µL.

30 Select the program **IndiMag\_Pathogen\_KF\_Duo\_4wash** on the instrument.


31 Start the run, then load the prepared plate and Elution Strip into position when prompted by the instrument.

#### Note

Keep the door open while extraction is in process. The chamber of the KingFisher Duo prime is small. Closing the door makes the ethanol vapor restrained inside the chamber and increases the ethanol contamination.

## 9. QUANTIFICATION AND STORAGE

32 After the running protocol is completed (~35 minutes), immediately remove the elution plate from the instrument and cover the plate or transfer the eluate to the final tube or plate of choice for final storage.

33 In a 0.6-mL microcentrifuge tube, use  1 µL total nucleic acid for DNA and RNA concentration measurement using Qubit 4 Fluorometer following manufacturer instructions.



Note

**Kits needed:** Qubit 1X dsDNA HS Assay Kit and Qubit RNA HS Assay Kit) (see Appendix 2 and Appendix 3.

**NOTE:** If the DNA or RNA concentration measurement is too high, dilute 2 µL of sample in a new 1.5 mL tube with 18 µL nuclease-free water. Use 3 µL of the diluted sample for DNA or RNA quantification.

34 Proceed with sample testing following the REDI-NET SOP F-4 Fecal Testing or store at -20 °C for less than 2 weeks.

Note

For long-term storage the sample needs to be stored at -80 °C following the REDI-NET SOP F-3 Fecal Storage.

APPENDIX 2. DNA and RNA Measurement using QUBIT FLUOROMETER 4.0

35 DNA quantification: 2m  
According to the volume of sample used, add the 1x HS dsDNA Qubit Assay for a final volume of 200 µL (i.e., if using 1 µL of sample, add 199 µL of 1x HS dsDNA Qubit Assay). Vortex for 5 - 10 seconds, then incubate for 00:02:00 at Room temperature before reading.






36 RNA Quantification:

36.1 In a new microcentrifuge tube/falcon tube (depending on the number of samples processed), prepare a working solution of the Qubit HS RNA Assay:

A	B	C
Reagents	Volume/sample	Volume for n+1 sample
Qubit RNA HS Assay buffer	199 µL	.... µL
Qubit RNA HS Assay Dye	1 µL	.... µL

36.2



In a new 0.6 ml tube, mix  199  $\mu\text{L}$  of Qubit HS RNA Assay working solution and  1  $\mu\text{L}$  of  2m the sample. Vortex for 5 - 10 seconds, then incubate for  00:02:00 at  Room temperature before reading.