

ILLUSTRATIVE EXEMPLAR — “what good looks like”

510(k) Cybersecurity Section — Summary

Connected infusion system (eSTAR)

Grounded in (public record): BD Alaris™ Infusion System (510(k) cleared July 2023). Public source: BD press release, news.bd.com (2023-07-21).

Regulatory basis: FD&C; Act §524B; FDA Premarket Cybersecurity Guidance (2025); eSTAR.

Section contents

#	Item	Reference artifact
1	Cyber-device determination & connectivity	CSMP §1
2	Security risk management summary	SRM report (SW96/14971)
3	Threat model summary	Threat model (STRIDE) + DFDs
4	Architecture views	Architecture package (4 views)
5	Security controls (8 categories)	Controls matrix + design evidence
6	SBOM	CycloneDX 1.5 + known-vuln list
7	Cybersecurity testing	SAST/DAST/SCA/fuzz/pen reports
8	Postmarket management + CVD	Postmarket plan + CVD policy
9	Security labeling	MDS2 + hardening guide
10	Reasonable-assurance conclusion	Summary statement

Conclusion

Each artifact above is summarized in the section and attached to the eSTAR submission; the section concludes with an explicit statement of **reasonable assurance of cybersecurity**, the standard FDA reviewers look for in a cyber-device 510(k).