

Postmarket Cybersecurity Plan

Grounded in: Abbott FreeStyle Libre (connected CGM)

Basis: FDA Postmarket guidance; AAMI TIR97.

Device	Abbott FreeStyle Libre (connected CGM)
Basis	FDA Postmarket Cybersecurity guidance; §524B(b)(1); AAMI TIR97. Illustrative.

1. Monitoring

- SBOM→CVE matching, CISA KEV, mobile-platform advisories (iOS/Android), researcher reports via CVD.

2. Assessment & remediation

Exploitability × patient-harm scoring (glucose-value integrity prioritized); severity-based app/firmware update SLAs; staged rollout via app stores with rollback.

3. Reporting & metrics

Metric	Example target
Critical patch (app/firmware)	≤ 30 days
% users on supported app version	≥ 95%
Open critical vulnerabilities	0

Sources (public)

- Abbott — FreeStyle Libre 2 Plus / 3 Plus CGM; FDA-cleared for use with automated insulin delivery (Mar 2023).
- Public connectivity facts: Bluetooth to a mobile app; readings transmitted to phone; AID integration.
- FDA — Cybersecurity in Medical Devices premarket guidance (2025); OWASP MASVS.