

Premarket Cybersecurity Documentation Checklist

Grounded in: BD Alaris™ Infusion System

Basis: FDA §524B acceptance criteria.

Device	BD Alaris™ Infusion System (connected Class II)
Use	Pre-submission acceptance gate — every item linked to an artifact in the eSTAR.

Acceptance checklist (all satisfied)

Item (§524B / 2025 guidance)	Artifact reference
Cyber-device determination & rationale	CSMP §1 (Wi-Fi + EMR interoperability)
Security risk management plan + file	SRM report (SW96/14971)
Threat model + data-flow diagrams	Threat model (STRIDE)
Security architecture views (4)	Architecture package
Security controls matrix (8 categories)	Controls matrix
SBOM (machine + human readable) + known vulns	CycloneDX SBOM + vuln list
Cybersecurity testing (SAST/DAST/SCA/fuzz/pen)	Test reports + coverage rationale
Postmarket vulnerability management + CVD plan	Postmarket + CVD plan
Security labeling / MDS2 / hardening guide	Labeling + MDS2
Interoperability + misconnection risk analysis	Interoperability analysis

Why this matters

Since October 1, 2023, FDA refuses to accept (RTA) cyber-device submissions missing §524B content. Each unchecked box above is a likely RTA; attaching the named artifact next to each item is what an excellent submission does.

Sources (public)

- FDA / BD — “BD Receives FDA 510(k) Clearance for Updated BD Alaris Infusion System” (Jul 21, 2023).
- FDA — Cybersecurity in Medical Devices: QMS Considerations and Content of Premarket Submissions (final, Jun 27, 2025).
- AAMI SW96; ISO 14971; IEC 81001-5-1.