

Cybersecurity Management Plan

Grounded in: Abbott FreeStyle Libre 3 Plus (connected CGM)

Basis: FDA 2025 guidance structure; device facts public.

Manufacturer	Abbott
Device	FreeStyle Libre 3 Plus — connected CGM
Connectivity	Bluetooth → mobile app (iOS/Android); cleared for AID integration
Basis	FDA Premarket Cybersecurity guidance (2025); SW96; ISO 14971. Illustrative.

1. Scope & context

Wearable CGM streaming readings over BLE to a mobile app and, when integrated, to an automated insulin delivery (AID) system. Related systems: mobile app, cloud, AID controller.

2. Security risk management

SW96 + ISO 14971 security risk file; emphasis on integrity of glucose values feeding dosing decisions; residual risk accepted by Security + Quality.

3. Controls (8 categories)

Category	Representative control
Authentication	Bonded BLE pairing to the app
Integrity	Validated/signed readings feeding AID
Cryptography	TLS to cloud; protected mobile storage
Confidentiality	PHI minimization in app + cloud
Updates	Signed app/sensor firmware updates

4. Postmarket & CVD

SBOM→CVE monitoring; CVD policy; severity-based patch timelines; coordination with AID partners (e.g., pump makers).

Sources (public)

- Abbott — FreeStyle Libre 2 Plus / 3 Plus CGM; FDA-cleared for use with automated insulin delivery (Mar 2023).
- Public connectivity facts: Bluetooth to a mobile app; readings transmitted to phone; AID integration.
- FDA — Cybersecurity in Medical Devices premarket guidance (2025); OWASP MASVS.