

Penetration Test Summary

Grounded in: Dexcom-class connected CGM (illustrative)

Basis: Threat-model-driven; OWASP MASVS.

Device	Dexcom-class connected CGM (illustrative)
Tester / scope	Independent accredited lab — sensor/BLE, mobile app (iOS/Android), cloud APIs
Basis	Threat-model-driven; OWASP MASVS/MASTG + API + BLE testing.

Findings

ID	Severity	Finding	Status
P-1	Medium	Verbose API error leakage	Closed
P-2	Low	Missing cert-pinning on one endpoint	Closed
P-3	Info	Logging hygiene	Risk-accepted

Coverage & conclusion

BLE, mobile storage/transport, API authorization, and cloud IAM mapped to the threat model. No high/critical open findings; residual risk acceptable; retest evidence attached.

Sources (public)

- FDA / Dexcom — Dexcom G7 510(k) clearance (Dec 2022); G6 first interoperable iCGM (2018, special controls).
- Public connectivity facts: Bluetooth Low Energy to the G7 app (iOS/Android) and Apple Watch; Dexcom Clarity cloud.
- FDA — Cybersecurity in Medical Devices premarket guidance (2025); OWASP MASVS for mobile.