FREEZER IOT TELEMETRY LOG
CONFIDENTIAL // FORENSIC INVESTIGATION RECORD

HELIOS BIOMEDICAL - CRYOGENIC FREEZER TELEMETRY (UNIT 4)
2024-08-14 01:45:00 - Temp: -80.2 C - Status: NORMAL
2024-08-14 02:10:00 - POWER FAULT: Mains disconnected - Generator: OFFLINE
2024-08-14 04:30:00 - Temp: -22.5 C - CRITICAL WARMING ALARM
2024-08-14 06:15:00 - Temp: +4.1 C - PRODUCT VIABILITY COMPROMISED
2024-08-14 07:45:00 - Mains power restored - Compressor active
2024-08-14 11:00:00 - Temp: -78.9 C - Status: RECOVERED