

Report No. 24PYT3\_DECPM12240024

Page 1 of 3

# **Preventive Maintenance Action Report**

	COM VONVAEW	DUVADIN TONCULIANC
Performed by :	Gen	Approved by :
ISSUE DATE	: 30 DECEM	BER 2024
PM. DATE	: 1 DECEMB	BER 2024
	Tel: (662) 4	467-1111 Fax: (622) 467-1111
	111 Phet Kasen	m Rd., Pak Khlong Phasi Charoen, Phasi Charoen Bangkok 1016
SUBMITTED BY	: РНҮАТНА	I 3 HOSPITAL
LOCATION	: WARD 11	
SERIAL No.	: 360125-M2	1505660005
MODEL	: M3A	
MANUFACTURE	R : EDAN	
ID CODE	: PYT3_0795	50
EQUIPMENT	: MONITORS	S, PHYSIOLOGIC, VITAL SIGNS

#### **CONDITION OF THIS RESULT OF TEST**

1. REFERENCE STANDARD INSTRUMENT:

MASTER MANUFACTURER MODEL SERIAL NO. CERTIFICATE NO. DUE DATE

2. THIS CERTIFICATION IS TRACEABLE TO:

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



*Report No.* 24PYT3\_DECPM12240024

Page 2 of 3

# **Preventive Maintenance Action Report**

EQUIPMENT : MONITORS, PHYSIOLOGIC, VITAL SIGNS

ID CODE : PYT3\_07950

SERIAL No. : 360125-M21505660005

### 1. Qualitative Tests

No.	List	Pass	Fail	None	Comment
1	Chassis/Housing	/			
2	Mount/Fasteners	/			
3	Casters/Brakes	/			
4	AC Plug	/			
5	Line Cord	/			
6	Strain Reliefs	/			
7	Circuit Breaker/Fuse	/			
8	Tubes/Bulbs	/			
9	Fittings/Connectors	/			
10	Transducers	/			
11	Controls/Switches	/			
12	Battery/Charger	/			
13	Time/Date Settings	/			
14	Cybersecurity			/	
15	User Calibration/Self-Test			/	
16	Alarms	/			
17	Labeling	/			
18	Accessories	/			

## 2. Quantitative Tests

No.	List	Pass	Fail	None	Comment
1	Grounding Resistance ≤0.5 Ω	/			0.172 Ω
2	Chassis Leakage Current ≤500 µA	/			2.7 μΑ
3	Air Leakage ≤15 mmHg/min	/			
4	Static Pressure Accuracy ±3 mmHg or	/			
	*Oscillometric Pressure ±8 mmHg				
5	Maximum Pressure (NIBP) ≤330 mmHg	/			
6	Heart Rate ±10%	/			

#### 3. Preventive Maintenance

No.	List	Pass	Fail	None	Comment
1	Clean the exterior and accessories	/			
2	Lubricate			/	



Report No. 24PYT3\_DECPM12240024

Page 3 of 3

# **Preventive Maintenance Action Report**

EQUIPMENT : MONITORS, PHYSIOLOGIC, VITAL SIGNS

ID CODE : PYT3\_07950

SERIAL No. : 360125-M21505660005

### 3. Preventive Maintenance (Continues)

No.	List	Pass	Fail	None	Comment
3	Adjust			/	
4	Replace			/	

FI-BME-NHS-CP-007/3 Rev.04 Page 3/3 Issued Date 20/07/2024