

Report No. 24PYT3_DECPM12240130

Page 1 of 2

Preventive Maintenance Action Report

EQUIPMENT : ASPIRATORS

ID CODE : PYT3_09437

MANUFACTURER : YUYUE

MODEL 7E-A

SERIAL No. 00081

LOCATION : BME-PYT3

SUBMITTED BY : PHYATHAI 3 HOSPITAL

111 Phet Kasem Rd., Pak Khlong Phasi Charoen, Phasi Charoen Bangkok 10160

Tel: (662) 467-1111 Fax: (622) 467-1111

PM. DATE : 4 DECEMBER 2024

ISSUE DATE : 30 DECEMBER 2024

Performed by:

PUKARIN TONGKLIANG

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTER MANUFACTURER MODEL SERIAL NO. CERTIFICATE NO. DUE DATE

2. THIS CERTIFICATION IS TRACEABLE TO:

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF **TEST ONLY**



Report No. 24PYT3_DECPM12240130

Page 2 of 2

Preventive Maintenance Action Report

EQUIPMENT : ASPIRATORS

ID CODE : PYT3_09437

SERIAL No. : 00081

1. Qualitative Tests

No.	List	Pass	Fail	None	Comment
1	Chassis/Housing	/			
2	Tubes/Hoses	/			
3	Fittings/Connectors	/			
4	Filters			/	
5	Controls/Switches	/			
6	Overflow Protection			/	
7	Intermittent Operation			/	
8	Indicators/Display	/			
9	Labeling	/			
10	Accessories	/			

2. Quantitative Tests

No.	List	Pass	Fail	None	Comment
1	Maximum Free Flow surgical/tracheal ≥25 L/min,			/	
	thoracic ≥20 L/min				
2	Maximum Vacuum Low-volume>100 mmHg;			/	
	thoracic >50 mmHg; surgical/tracheal ≥400 mmHg				
3	Vacuum Gauge Accuracy ±5% FSO(Full Scale	/			
	output)				

3. Preventive Maintenance

No.	List	Pass	Fail	None	Comment
1	Clean the exterior and accessories	/			
2	Lubricate			/	
3	Adjust			/	
4	Replace			/	