

Certificate No.

24PYT3_DEC12240133

Page 1 of 2

Certificate of Calibration

EQUIPMENT : ASPIRATORS, EMERGENCY

ID CODE : PYT3_08606

MANUFACTURER : BOSCAROL

MODEL : OB1000

SERIAL No. : 2222110417

LOCATION : AMBULANCE (รถพยาบาล)

SUBMITTED BY : PHYATHAI 3 HOSPITAL

111 Phet Kasem Rd., Pak Khlong Phasi Charoen, Phasi Charoen Bangkok 10160

Tel: (662) 467-1111 Fax: (622) 467-1111

CALIBRATED DATE : 4 DECEMBER 2024

ISSUE DATE : 30 DECEMBER 2024

Performed by :		Approved by :	1	
	SOM KONKA FW		PUKARIN TONGKUANG	

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATETESTER,FLUKE717-30G59750045L202404137-000130 Apr 2025

PRESSURE CALIBRATOR

2. THIS CERTIFICATION IS TRACEABLE TO:

- Temperature, Dimension, mass, Anemometer, Electrical/Frequency, Relative Humidity

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Certificate No. 24PYT3_DEC12240133

Page 2 of 2

Calibration Report

EQUIPMENT : ASPIRATORS, EMERGENCY

ID CODE : PYT3_08606

MANUFACTURER : BOSCAROL

MODEL : OB1000

SERIAL No. : 2222110417

DATE OF CALIBRATION : 4 DECEMBER 2024

ENVIRONMENT : TEMPERATURE 25 °C

RELATIVE HUMIDITY 55 %

PROCEDURE USED:

This instrument was calibration by comparison with standard

MEASURMENT RESULT:

/	Without Adjustment		Before Adjustment		After Adjustment
---	--------------------	--	-------------------	--	------------------

VACUUM				
UUC* Setting	Standard Reading	Error	% Error	Uncertainty
(mmHg)	(mmHg)	(mmHg)	(%)	(mmHg)
150	151.15	1.15	0.77	± 0.0547
300	307.21	7.21	2.40	± 0.0826
450	454.19	4.19	0.93	± 0.0776
600	604.03	4.03	0.67	± 0.0901
750	740.40	-9.60	-1.28	± 0.1025

UUC*: Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

2301/2 New Petchburi Soi 47 (soonvijai), Bangkapi, Huaykwang, Bangkok 10310 Tel: (66) 2762 4000 Fax: (66) 2762 4072

FI-BME-NHS-CP-015/1 Rev.04 Page 2/2 Issued Date 20/07/2024