

MODEL

Certificate No.

24PYT3_DEC12240124

Page 1 of 2

Certificate of Calibration

EQUIPMENT : OXIMETERS, PULSE

ID CODE : PYT3_00604

MANUFACTURER : EDAN

SERIAL No. : 10210A15539

LOCATION : AMBULANCE (รถพยาบาล) (ห้องเก็บของ)

H100B

SUBMITTED BY : PHYATHAI 3 HOSPITAL

111 Phet Kasem Rd., Pak Khlong Phasi Charoen, Phasi Charoen Bangkok 10160

Tel: (662) 467-1111 Fax: (622) 467-1111

CALIBRATED DATE : 4 DECEMBER 2024
ISSUE DATE : 30 DECEMBER 2024

Performed by:		Approved by :			
	SOM KONKAEW		PUKARIN TONGKLIANG		

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATETester, Vital SignFLUKEPROSIM 8570205924MD54530 Apr 2025

Simulator

2. THIS CERTIFICATION IS TRACEABLE TO:

- Technology Promotion Association (Thai-Japan)

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Certificate No. 24PYT3_DEC12240124

Page 2 of 2

Calibration Report

EQUIPMENT : OXIMETERS, PULSE

ID CODE : PYT3_00604

MANUFACTURER : EDAN MODEL : H100B

SERIAL No. : 10210A15539

DATE OF CALIBRATION : 4 DECEMBER 2024

ENVIRONMENT : TEMPERATURE 25 °C

RELATIVE HUMIDITY 55 %

PROCEDURE USED:

This instrument was calibration by comparison with standard

MEASURMENT RESULT:

/	Without Adjustment		Before Adjustment		After Adjustment
---	--------------------	--	-------------------	--	------------------

%Spo2				
Standard Setting	UUC* Reading	Error	% Error	Uncertainty
(%SPO ₂)	(%SPO ₂)	(%SPO ₂)	(%)	(%SPO ₂)
90	90.99	0.99	1.10	± 0.0752
96	97.01	1.01	1.05	± 0.0579
100	100.01	0.01	0.01	± 0.0799

Heart Rate						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(BPM)	(BPM)	(BPM)	(%)	(BPM)		
60	60.05	0.05	0.08	± 0.0487		
80	80.04	0.03	0.04	± 0.1194		
120	119.99	-0.01	-0.01	± 0.1139		

UUC*: Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%

FI-BME-NHS-CP-012/1 Rev.04 Page 2/2 Issued Date 20/07/2024