

Report No. 24PYT3\_DECPM12240127

Page 1 of 3

# **Preventive Maintenance Action Report**

MONITORS, PATIENT TRANSPORT **EQUIPMENT** 

**ID CODE** PYT3\_02752

MANUFACTURER **PHILIPS** 

**MODEL SURESIGNS VSI** 

SERIAL No. CN21208242

**LOCATION** : PHYSICAL THERAPY( กายภาพ)

SUBMITTED BY : PHYATHAI 3 HOSPITAL

111 Phet Kasem Rd., Pak Khlong Phasi Charoen, Phasi Charoen Bangkok 10160

Tel: (662) 467-1111 Fax: (622) 467-1111

PM. DATE 4 DECEMBER 2024 **ISSUE DATE** : 30 DECEMBER 2024

Performed by: Approved by: PUKARIN TONGKLIANG

#### **CONDITION OF THIS RESULT OF TEST**

#### 1. REFERENCE STANDARD INSTRUMENT:

MASTER MANUFACTURER MODEL SERIAL NO. CERTIFICATE NO. DUE DATE **FLUKE** ESA614 2421039 ELECTRICAL L202403400-0007 30 May 2025

**SAFETY ANALYZER** 

#### 2. THIS CERTIFICATION IS TRACEABLE TO:

- Temperature, Dimension, mass, Anemometer, Electrical/Frequency, Relative Humidity

## 3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF **TEST ONLY**



*Report No.* 24PYT3\_DECPM12240127

Page 2 of 3

## **Preventive Maintenance Action Report**

EQUIPMENT : MONITORS, PATIENT TRANSPORT

ID CODE : PYT3\_02752 SERIAL No. : CN21208242

## 1. Qualitative Tests

No.	List	Pass	Fail	None	Comment
1	Chassis/Housing	/			
2	Mount/Fasteners	/			
3	Casters/Brakes	/			
4	AC Plug	/			
5	Line Cord	/			
6	Strain Reliefs	/			
7	Circuit Breaker/Fuse	/			
8	Tubes/Bulbs	/			
9	Fittings/Connectors	/			
10	Transducers	/			
11	Controls/Switches	/			
12	Battery/Charger	/			
13	Time/Date Settings	/			
14	Cybersecurity			/	
15	User Calibration/Self-Test			/	
16	Alarms	/			
17	Labeling	/			
18	Accessories	/			

### 2. Quantitative Tests

No.	List	Pass	Fail	None	Comment
1	Grounding Resistance ≤0.5 Ω	/			0.461 Ω
2	Chassis Leakage Current ≤500 μA	/			135.6 μΑ
3	Air Leakage ≤15 mmHg/min	/			
4	Static Pressure Accuracy ±3 mmHg or	/			
	*Oscillometric Pressure ±8 mmHg				
5	Maximum Pressure (NIBP) ≤330 mmHg	/			
6	Heart Rate ±10%	/			

#### 3. Preventive Maintenance

No.	List	Pass	Fail	None	Comment
1	Clean the exterior and accessories	/			
2	Lubricate			/	



*Report No.* 24PYT3\_DECPM12240127

Page 3 of 3

## **Preventive Maintenance Action Report**

EQUIPMENT : MONITORS, PATIENT TRANSPORT

ID CODE : PYT3\_02752 SERIAL No. : CN21208242

### 3. Preventive Maintenance (Continues)

No.	List	Pass	Fail	None	Comment
3	Adjust			/	
4	Replace			/	

FI-BME-NHS-CP-007/3 Rev.04 Page 3/3 Issued Date 20/07/2024