

Certificate No.

24PYT3_DEC12240039

Page 1 of 2

Certificate of Calibration

EQUIPMENT : OXIMETERS, PULSE

ID CODE : PYT3_07566

MANUFACTURER : EDAN

MODEL : H100B

SERIAL No. : 360101-M20306580091

LOCATION : WARD 10

SUBMITTED BY : PHYATHAI 3 HOSPITAL

111 Phet Kasem Rd., Pak Khlong Phasi Charoen, Phasi Charoen Bangkok 10160

Tel: (662) 467-1111 Fax: (622) 467-1111

CALIBRATED DATE : 1 DECEMBER 2024

ISSUE DATE : 30 DECEMBER 2024

Performed by : Approved by : PUKARIN TONGKLIANG

SOM KONKAEW PUKARIN TONGKLIANG

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATETester, Vital SignFLUKEPROSIM 8570205924MD54530 Apr 2025Simulator

2. THIS CERTIFICATION IS TRACEABLE TO:

- Technology Promotion Association (Thai-Japan)

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Certificate No. 24PYT3_DEC12240039

Page 2 of 2

Calibration Report

EQUIPMENT : OXIMETERS, PULSE

ID CODE : PYT3_07566

MANUFACTURER : EDAN MODEL : H100B

SERIAL No. : 360101-M20306580091

DATE OF CALIBRATION : 1 DECEMBER 2024

ENVIRONMENT : TEMPERATURE 25 °C

RELATIVE HUMIDITY 55 %

PROCEDURE USED:

This instrument was calibration by comparison with standard

MEASURMENT RESULT:

/	Without Adjustment		Before Adjustment		After Adjustment
---	--------------------	--	-------------------	--	------------------

%Spo2								
Standard Setting	UUC* Reading	Error	% Error	Uncertainty				
(%SPO ₂)	(%SPO ₂)	(%SPO ₂)	(%)	(%SPO ₂)				
90	91.02	1.02	1.13	± 0.0283				
96	97.00	1.00	1.04	± 0.1081				
100	100.01	0.01	0.01	± 0.1024				

Heart Rate									
Standard Setting	UUC* Reading	Error	% Error	Uncertainty					
(BPM)	(BPM)	(BPM)	(%)	(BPM)					
60	59.99	-0.01	-0.02	± 0.0600					
80	81.01	1.01	1.26	± 0.1153					
120	120.03	0.03	0.03	± 0.1140					

UUC*: Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

FI-BME-NHS-CP-012/1 Rev.04 Page 2/2 Issued Date 20/07/2024