

Report No. 24PYT3_DECPM12240137

Page 1 of 2

10160

Preventive Maintenance Action Report

EQUIPMENT	:	OXIMETERS, PULSE
ID CODE	:	PYT3_05601
MANUFACTURER	:	NONIN
MODEL	:	9500 (ONYX)
SERIAL No.	:	502483447
LOCATION	:	WARD 17
SUBMITTED BY	:	PHYATHAI 3 HOSPITAL
		111 Phet Kasem Rd., Pak Khlong Phasi Charoen, Phasi Charoen Bangkok

Tel: (662) 467-1111 Fax: (622) 467-1111

PM. DATE : 1 DECEMBER 2024 ISSUE DATE : 30 DECEMBER 2024

nen

Performed by:	7	Approved by :	2-7-		
	SOM KONKAFW		PUKARIN TONGKLIANG		

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTER MANUFACTURER MODEL SERIAL NO. CERTIFICATE NO. DUE DATE

2. THIS CERTIFICATION IS TRACEABLE TO:

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF **TEST ONLY**



Report No. 24PYT3_DECPM12240137

Page 2 of 2

Preventive Maintenance Action Report

EQUIPMENT : OXIMETERS, PULSE

ID CODE : PYT3_05601

SERIAL No. : 502483447

1. Qualitative Tests

No.	List	Pass	Fail	None	Comment
1	Chassis/Housing	/			
2	Mount/Fasteners			/	
3	AC Plug/Receptacles			/	
4	Line Cord			/	
5	Strain Reliefs			/	
6	Circuit Breaker/Fuse			/	
7	Cables			/	
8	Connectors	/			
9	Probes/Sensors/Transducers	/			
10	Controls/Switches	/			
11	Battery/Charger	/			
12	Indicators/Displays	/			
13	Self-Test	/			
14	Time/Date Settings	/			
15	Cybersecurity			/	
16	Alarms	/			
17	Audible Signals	/			
18	Labeling	/			

2. Quantitative Tests

No.	List	Pass	Fail	None	Comment
1	Grounding Resistance ≤0.5 Ω			/	
2	Chassis Leakage Current ≤500 µA			/	

3. Preventive Maintenance

No.	List	Pass	Fail	None	Comment
1	Clean the exterior and accessories	/			
2	Lubricate			/	
3	Adjust			/	
4	Replace			/	