

Certificate No.

24PYT3_DEC12240026

Page 1 of 2

Certificate of Calibration

EQUIPMENT : OXIMETERS, PULSE

ID CODE : PYT3_04701

MANUFACTURER : EDAN

MODEL : H100B

SERIAL No. : 113575-M15500140008

LOCATION : WARD 11

SUBMITTED BY : PHYATHAI 3 HOSPITAL

111 Phet Kasem Rd., Pak Khlong Phasi Charoen, Phasi Charoen Bangkok 10160

Tel: (662) 467-1111 Fax: (622) 467-1111

CALIBRATED DATE : 1 DECEMBER 2024

ISSUE DATE : 30 DECEMBER 2024

Performed by : Approved by : PUKARIN TONGKLIANG

SOM KONKAEW PUKARIN TONGKLIANG

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATETester, Vital SignFLUKEPROSIM 8570205924MD54530 Apr 2025

Simulator

2. THIS CERTIFICATION IS TRACEABLE TO:

- Technology Promotion Association (Thai-Japan)

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Certificate No. 24PYT3_DEC12240026

Page 2 of 2

Calibration Report

EQUIPMENT : OXIMETERS, PULSE

ID CODE : PYT3_04701

MANUFACTURER : EDAN

SERIAL No. : 113575-M15500140008

DATE OF CALIBRATION : 1 DECEMBER 2024

ENVIRONMENT : TEMPERATURE 25 °C

: H100B

RELATIVE HUMIDITY 55 %

PROCEDURE USED:

MODEL

This instrument was calibration by comparison with standard

MEASURMENT RESULT:

/	Without Adjustment		Before Adjustment		After Adjustment
---	--------------------	--	-------------------	--	------------------

%Spo2								
Standard Setting	UUC* Reading	Error	% Error	Uncertainty				
(%SPO ₂)	(%SPO ₂)	(%SPO ₂)	(%)	(%SPO ₂)				
90	90.75	0.75	0.83	± 0.8275				
96	96.00	-0.01	-0.01	± 0.1025				
100	99.99	-0.01	-0.01	± 0.0248				

Heart Rate								
Standard Setting	UUC* Reading	Error	% Error	Uncertainty				
(BPM)	(BPM)	(BPM)	(%)	(BPM)				
60	60.01	0.01	0.02	± 0.1001				
80	80.00	0.00	0.00	± 0.1113				
120	119.99	-0.02	-0.01	± 0.1152				

UUC*: Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

FI-BME-NHS-CP-012/1 Rev.04 Page 2/2 Issued Date 20/07/2024