

Certificate No.

24PTP_0104240400

Page 1 of 2

Certificate of Calibration

EQUIPMENT : INFUSION PUMPS, ENTERAL FEEDING

ID CODE : PTP04602

MANUFACTURER : HAWKMED

MODEL : HK-300

SERIAL No. : 01163190

LOCATION : I.C.U

SUBMITTED BY : PHYATHAI PHAHOLYOTHIN HOSPITAL

670/1 Phaholyothin Road, Samsen Nai, Phaya Thai, Bangkok 10400

Tel: 02-271-7000 Fax: -

CALIBRATED DATE : 30 APRIL 2024

ISSUE DATE : 3 MAY 2024

Supatcha

Performed by : Approved by : SUPATCHA BUAPUEAN

d by :

PHUWASIN YIWKIM

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATEInfusion PumpRIGEL MedicalMulti-Flo23Q-141223MD116924 Aug 2024

Tester

writing from the calibration.

2. THIS CERTIFICATION IS TRACEABLE TO:

- Technology Promotion Association (Thai-Japan)

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Certificate No. 24PTP_0104240400

Page 2 of 2

Calibration Report

			cport	
EQUIPMENT		NFUSION PUMPS, EN	TERAL FEEDING	
ID CODE	: P'	TP04602		
MANUFACTURER	: Н	[AWKMED		
MODEL		IK-300		
SERIAL No.				
SERIAL NO.	: 0	1163190		
DATE OF CALIBRA	ATION : 30	0 APRIL 2024		
ENVIRONMENT	: T	EMPERATURE	24 °C	
	R	ELATIVE HUMIDITY	55 %	
/ Without Adjus	tment	Before Adjustment	After A	djustment
UUC* Setting	Standard Reading	Error	% Error	Uncertaint
(ml/hr)	(ml/hr)	(ml/hr)	(%)	(ml/hr)
100	107.71	7.71	7.71	± 0.2601
200	214.97	14.97	7.48	± 0.7101
300	322.17	22.17	7.39	± 0.7102
	ME			
Use IV Set Brand:				

FI-BME-NHS-CP-072/1 Rev.03 Page 2/2 Issued Date 20/07/2023