

Certificate No. 24

24PTP\_0104240299

Page 1 of 3

### **Certificate of Calibration**

EQUIPMENT : MONITORS, BEDSIDE, PHYSIOLOGIC

ID CODE : PTP00697

MANUFACTURER : NIHON KOHDEN

MODEL : BSM-2303K

SERIAL No. : 1369

LOCATION : Anesthesia

SUBMITTED BY : PHYATHAI PHAHOLYOTHIN HOSPITAL

670/1 Phaholyothin Road, Samsen Nai, Phaya Thai, Bangkok 10400

Tel: 02-271-7000 Fax: -

CALIBRATED DATE : 24 APRIL 2024

USANEE CHALONGPHOKSILCHAI

ISSUE DATE : 3 MAY 2024

Performed by:

Approved by:

PHUWASIN YIWKIM

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

### **CONDITION OF THIS RESULT OF TEST**

#### 1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATEVitalsign SimulatorFlukeProSim8254103123MD7777 Jun 2024

#### 2. THIS CERTIFICATION IS TRACEABLE TO:

- Technology Promotion Association (Thai-Japan)

## 3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Certificate No. 24PTP\_0104240299

Page 2 of 3

# **Calibration Report**

EQUIPMENT : MONITORS, BEDSIDE, PHYSIOLOGIC

ID CODE : PTP00697

MANUFACTURER : NIHON KOHDEN

MODEL : BSM-2303K

SERIAL No. : 1369

DATE OF CALIBRATION : 24 APRIL 2024

ENVIRONMENT : TEMPERATURE 22 °C

RELATIVE HUMIDITY 55 %

### **PROCEDURE USED:**

This instrument was calibration by comparison with standard

### **MEASURMENT RESULT:**

/	Without Adjustment		Before Adjustment		After Adjustment
---	--------------------	--	-------------------	--	------------------

Systolic(NIBP)						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(mmHg)	(mmHg)	(mmHg)	(%)	(mmHg)		
60	62.00	2.00	3.33	± 0.9238		
120	118.00	-2.00	-1.67	± 1.2702		
200	199.00	-1.00	-0.50	± 1.7321		

Diastolic(NIBP)					
Standard Setting	UUC* Reading	Error	% Error	Uncertainty	
(mmHg)	(mmHg)	(mmHg)	(%)	(mmHg)	
30	32.00	2.00	6.67	± 0.7506	
80	83.00	3.00	3.75	± 1.0393	
150	152.00	2.00	1.33	± 1.4434	

Heart Rate						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(BPM)	(BPM)	(BPM)	(%)	(BPM)		
60	60.00	0.00	0.00	± 0.0150		
80	80.00	0.00	0.00	± 0.0460		
120	120.00	0.00	0.00	± 0.0460		

FI-BME-NHS-CP-029/1 Rev.03 Page 2/3 Issued Date 20/07/2023



Certificate No. 24PTP\_0104240299

Page 3 of 3

# **Calibration Report**

EQUIPMENT : MONITORS, BEDSIDE, PHYSIOLOGIC

ID CODE : PTP00697

SERIAL No. : 1369

% Spo2						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(%SPO <sub>2</sub> )	(%SPO <sub>2</sub> )	(%SPO <sub>2</sub> )	(%)	(%SPO <sub>2</sub> )		
90	90.00	0.00	0.00	± 0.1039		
96	96.00	0.00	0.00	± 0.1109		
100	100.00	0.00	0.00	± 0.1155		

Temperature					
Standard Setting	UUC* Reading	Error	% Error	Uncertainty	
(°C)	(°C)	(°C)	(%)	(°C)	
35	35.00	0.00	0.00	± 0.0051	
37	37.10	0.10	0.27	± 0.0051	
39	39.00	0.00	0.00	± 0.0055	

### **UUC\*: Unit Under Calibration**

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%

FI-BME-NHS-CP-029/1 Rev.03 Page 3/3 Issued Date 20/07/2023