

Certificate No.

24PTP\_0104240393

Page 1 of 2

## **Certificate of Calibration**

EQUIPMENT : OXIMETERS, PULSE

ID CODE : PTP04035

MANUFACTURER : -

MODEL : -

SERIAL No. : -

LOCATION : BME-PTP

SUBMITTED BY : PHYATHAI PHAHOLYOTHIN HOSPITAL

670/1 Phaholyothin Road, Samsen Nai, Phaya Thai, Bangkok 10400

Tel: 02-271-7000 Fax: -

CALIBRATED DATE : 30 APRIL 2024

ISSUE DATE : 1 MAY 2024

Supatcha

3,1024

SUPATCHA BUAPUEAN PHUWASIN YIWKIM

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

Approved by:

### **CONDITION OF THIS RESULT OF TEST**

Performed by:

### 1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATEVitalsign SimulatorFlukeProSim8254103123MD7777 Jun 2024

#### 2. THIS CERTIFICATION IS TRACEABLE TO:

- Technology Promotion Association (Thai-Japan)

# 3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



*Certificate No.* 24PTP\_0104240393

Page 2 of 2

# **Calibration Report**

EQUIPMENT : OXIMETERS, PULSE

ID CODE : PTP04035

MANUFACTURER : -

MODEL : SERIAL No. :

DATE OF CALIBRATION : 30 APRIL 2024

ENVIRONMENT : TEMPERATURE 24 °C

RELATIVE HUMIDITY 56 %

#### **PROCEDURE USED:**

This instrument was calibration by comparison with standard

### **MEASURMENT RESULT:**

/	Without Adjustment		Before Adjustment		After Adjustment
---	--------------------	--	-------------------	--	------------------

%Spo2								
Standard Setting	UUC* Reading	Error	% Error	Uncertainty				
(%SPO <sub>2</sub> )	(%SPO <sub>2</sub> )	(%SPO <sub>2</sub> )	(%)	(%SPO <sub>2</sub> )				
90	90.00	0.00	0.00	± 0.1039				
96	96.00	0.00	0.00	± 0.1109				
100	99.00	-1.00	-1.00	± 0.1155				

Heart Rate									
Standard Setting	UUC* Reading	Error	% Error	Uncertainty					
(BPM)	(BPM)	(BPM)	(%)	(BPM)					
60	60.00	0.00	0.00	± 0.0150					
80	80.00	0.00	0.00	± 0.0460					
120	120.00	0.00	0.00	± 0.0460					

### **UUC\*: Unit Under Calibration**

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%

FI-BME-NHS-CP-012/1 Rev.03 Page 2/2 Issued Date 20/07/2023