

Certificate No.

24PTP_0104240288

Page 1 of 3

Certificate of Calibration

EQUIPMENT : MODULE, ECG/SPO2/NIBP/IBP/TEMP

ID CODE : PTP00626

MANUFACTURER : NIHON KOHDEN

MODEL : AY-663P

SERIAL No. : 6390

LOCATION : Operating Room

SUBMITTED BY : PHYATHAI PHAHOLYOTHIN HOSPITAL

670/1 Phaholyothin Road, Samsen Nai, Phaya Thai, Bangkok 10400

Tel: 02-271-7000 Fax: -

CALIBRATED DATE : 24 APRIL 2024

ISSUE DATE : 3 MAY 2024

Usanee

Approved by :

USANEE CHALONGPHOKSILCHAI

PHUWASIN YIWKIM

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

CONDITION OF THIS RESULT OF TEST

Performed by:

1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATEVitalsign SimulatorFlukeProSim8254103123MD7777 Jun 2024

2. THIS CERTIFICATION IS TRACEABLE TO:

- Technology Promotion Association (Thai-Japan)

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Certificate No. 24PTP_0104240288

Page 2 of 3

Calibration Report

EQUIPMENT : MODULE, ECG/SPO2/NIBP/IBP/TEMP

ID CODE : PTP00626

MANUFACTURER : NIHON KOHDEN

MODEL : AY-663P

SERIAL No. : 6390

DATE OF CALIBRATION : 24 APRIL 2024

ENVIRONMENT : TEMPERATURE 22 °C

RELATIVE HUMIDITY 55 %

PROCEDURE USED:

This instrument was calibration by comparison with standard

MEASURMENT RESULT:

/	Without Adjustment		Before Adjustment		After Adjustment
---	--------------------	--	-------------------	--	------------------

Systolic(NIBP)					
Standard Setting	UUC* Reading	Error	% Error	Uncertainty	
(mmHg)	(mmHg)	(mmHg)	(%)	(mmHg)	
60	63.00	3.00	5.00	± 0.9238	
120	121.00	1.00	0.83	± 1.2702	
200	202.00	2.00	1.00	± 1.7321	

Diastolic(NIBP)					
Standard Setting	UUC* Reading	Error	% Error	Uncertainty	
(mmHg)	(mmHg)	(mmHg)	(%)	(mmHg)	
30	32.00	2.00	6.67	± 0.7506	
80	82.00	2.00	2.50	± 1.0393	
150	152.00	2.00	1.33	± 1.4434	

Heart Rate					
Standard Setting	UUC* Reading	Error	% Error	Uncertainty	
(BPM)	(BPM)	(BPM)	(%)	(BPM)	
60	60.00	0.00	0.00	± 0.0150	
80	80.00	0.00	0.00	± 0.0460	
120	120.00	0.00	0.00	± 0.0460	

FI-BME-NHS-CP-029/1 Rev.03 Page 2/3 Issued Date 20/07/2023



Certificate No. 24PTP_0104240288

Page 3 of 3

Calibration Report

EQUIPMENT : MODULE, ECG/SPO2/NIBP/IBP/TEMP

ID CODE : PTP00626

SERIAL No. : 6390

%Spo2						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(%SPO ₂)	(%SPO ₂)	(%SPO ₂)	(%)	(%SPO ₂)		
90	90.00	0.00	0.00	± 0.1039		
96	96.00	0.00	0.00	± 0.1109		
100	100.00	0.00	0.00	± 0.1155		

Temperature					
Standard Setting	UUC* Reading	Error	% Error	Uncertainty	
(°C)	(°C)	(°C)	(%)	(°C)	
35	35.00	0.00	0.00	± 0.0051	
37	37.00	0.00	0.00	± 0.0051	
39	39.00	0.00	0.00	± 0.0055	

UUC*: Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%

FI-BME-NHS-CP-029/1 Rev.03 Page 3/3 Issued Date 20/07/2023