

Certificate No.

24PTP\_0104240084

Page 1 of 2

### **Certificate of Calibration**

EQUIPMENT : OXIMETERS, PULSE

ID CODE : PLP03866

MANUFACTURER : STRONG

MODEL : 703F

SERIAL No. : 703F-8007008

LOCATION : Physiotherapy

SUBMITTED BY : PHYATHAI PHAHOLYOTHIN HOSPITAL

670/1 Phaholyothin Road, Samsen Nai, Phaya Thai, Bangkok 10400

Tel: 02-271-7000 Fax: -

CALIBRATED DATE : 17 APRIL 2024

ISSUE DATE : 1 MAY 2024

Performed by:

Approved by :

USANEE CHALONGPHOKSILCHAI

PHUWASIN YIWKIM

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

### **CONDITION OF THIS RESULT OF TEST**

#### 1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATEVitalsign SimulatorFlukeProSim8254103123MD7777 Jun 2024

#### 2. THIS CERTIFICATION IS TRACEABLE TO:

- Technology Promotion Association (Thai-Japan)

## 3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Certificate No. 24PTP\_0104240084

Page 2 of 2

# **Calibration Report**

EQUIPMENT : OXIMETERS, PULSE

ID CODE : PLP03866

MANUFACTURER : STRONG

MODEL : 703F

SERIAL No. : 703F-8007008

DATE OF CALIBRATION : 17 APRIL 2024

ENVIRONMENT : TEMPERATURE 25 °C

RELATIVE HUMIDITY 55 %

### **PROCEDURE USED:**

This instrument was calibration by comparison with standard

### **MEASURMENT RESULT:**

/	Without Adjustment		Before Adjustment		After Adjustment
---	--------------------	--	-------------------	--	------------------

%Spo2								
Standard Setting	UUC* Reading	Error	% Error	Uncertainty				
(%SPO <sub>2</sub> )	(%SPO <sub>2</sub> )	(%SPO <sub>2</sub> )	(%)	(%SPO <sub>2</sub> )				
90	90.00	0.00	0.00	± 0.1039				
96	96.00	0.00	0.00	± 0.1109				
100	99.00	-1.00	-1.00	± 0.1155				

Heart Rate								
Standard Setting	UUC* Reading	Error	% Error	Uncertainty				
(BPM)	(BPM)	(BPM)	(%)	(BPM)				
60	60.00	0.00	0.00	± 0.0150				
80	81.00	1.00	1.25	± 0.0460				
120	120.00	0.00	0.00	± 0.0460				

### **UUC\*: Unit Under Calibration**

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%

FI-BME-NHS-CP-012/1 Rev.03 Page 2/2 Issued Date 20/07/2023