

*Report No.* 24PYT3\_DECPM12240048

PUKARIN TONGKLIANG

Page 1 of 2

# **Preventive Maintenance Action Report**

EQUIPMENT	: OXIMETERS, PULSE
ID CODE	: PYT3_00652
MANUFACTURER	: MASIMO
MODEL	: RADICAL-7
SERIAL No.	: 166731
LOCATION	: NURSERY(เด็กอ่อน)
SUBMITTED BY	: PHYATHAI 3 HOSPITAL
	111 Phet Kasem Rd., Pak Khlong Phasi Charoen, Phasi Charoen Bangkok 10160
	Tel: (662) 467-1111 Fax: (622) 467-1111
PM. DATE	: 1 DECEMBER 2024
ISSUE DATE	: 30 DECEMBER 2024
Performed by :	Approved by :

#### **CONDITION OF THIS RESULT OF TEST**

2. THIS CERTIFICATION IS TRACEABLE TO:

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF **TEST ONLY** 



Report No. 24PYT3\_DECPM12240048

Page 2 of 2

## **Preventive Maintenance Action Report**

EQUIPMENT : OXIMETERS, PULSE

ID CODE : PYT3\_00652

SERIAL No. : 166731

### 1. Qualitative Tests

No.	List	Pass	Fail	None	Comment
1	Chassis/Housing	/			
2	2 Mount/Fasteners			/	
3	AC Plug/Receptacles			/	
4	Line Cord			/	
5	Strain Reliefs			/	
6	Circuit Breaker/Fuse			/	
7	Cables			/	
8	Connectors	/			
9	Probes/Sensors/Transducers	/			
10	Controls/Switches	/			
11	Battery/Charger	/			
12	Indicators/Displays	/			
13	Self-Test	/			
14	Time/Date Settings	/			
15	Cybersecurity			/	
16	Alarms	/			
17	Audible Signals	/			
18	Labeling	/			

#### 2. Quantitative Tests

No.	List	Pass	Fail	None	Comment
1	Grounding Resistance ≤0.5 Ω			/	
2	Chassis Leakage Current ≤500 µA	/			201 μΑ

### 3. Preventive Maintenance

No.	List	Pass	Fail	None	Comment
1	Clean the exterior and accessories	/			
2	Lubricate			/	
3	Adjust			/	
4	Replace			/	