

Certificate No.

24PYT3\_DEC12240083

Page 1 of 3

### **Certificate of Calibration**

EQUIPMENT : MONITORS, PATIENT TRANSPORT

ID CODE : PYT3\_09309

MANUFACTURER : EDAN

MODEL : IM20

SERIAL No. : 260374-M24322910001

LOCATION : OPERATING ROOM

SUBMITTED BY : PHYATHAI 3 HOSPITAL

111 Phet Kasem Rd., Pak Khlong Phasi Charoen, Phasi Charoen Bangkok 10160

Tel: (662) 467-1111 Fax: (622) 467-1111

CALIBRATED DATE : 1 DECEMBER 2024

ISSUE DATE : 30 DECEMBER 2024

Performed by : Approved by : PUKARIN TONGKLIANG

SOM KONKAEW PUKARIN TONGKLIANG

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

### **CONDITION OF THIS RESULT OF TEST**

#### 1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATETester, Vital SignFLUKEPROSIM 8570205924MD54530 Apr 2025

Simulator

### 2. THIS CERTIFICATION IS TRACEABLE TO:

- Technology Promotion Association (Thai-Japan)

## 3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Certificate No. 24PYT3\_DEC12240083

Page 2 of 3

# **Calibration Report**

EQUIPMENT : MONITORS, PATIENT TRANSPORT

ID CODE : PYT3\_09309

MANUFACTURER : EDAN

MODEL : IM20

SERIAL No. : 260374-M24322910001

DATE OF CALIBRATION : 1 DECEMBER 2024

ENVIRONMENT : TEMPERATURE 25 °C

RELATIVE HUMIDITY 55 %

### **PROCEDURE USED:**

This instrument was calibration by comparison with standard

### **MEASURMENT RESULT:**

/	Without Adjustment		Before Adjustment		After Adjustment
---	--------------------	--	-------------------	--	------------------

Systolic						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(mmHg)	(mmHg)	(mmHg)	(%)	(mmHg)		
60	59.99	-0.01	-0.02	± 0.0563		
120	119.97	-0.03	-0.03	± 0.0589		
200	200.02	0.02	0.01	± 0.0475		

Diastolic						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(mmHg)	(mmHg)	(mmHg)	(%)	(mmHg)		
30	36.03	6.03	20.10	± 0.0405		
80	78.00	-2.00	-2.50	± 0.0880		
150	147.02	-2.99	-1.99	± 0.0770		

Heart Rate						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(BPM)	(BPM)	(BPM)	(%)	(BPM)		
60	59.95	-0.05	-0.08	± 0.0499		
80	80.01	0.01	0.01	± 0.1157		
120	120.02	0.02	0.01	± 0.1275		

FI-BME-NHS-CP-007/1 Rev.04 Page 2/3 Issued Date 20/07/2024



Certificate No. 24PYT3\_DEC12240083

Page 3 of 3

# **Calibration Report**

EQUIPMENT : MONITORS, PATIENT TRANSPORT

ID CODE : PYT3\_09309

SERIAL No. : 260374-M24322910001

%Spo2						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(%SPO <sub>2</sub> )	(%SPO <sub>2</sub> )	(%SPO <sub>2</sub> )	(%)	(%SPO <sub>2</sub> )		
90	90.98	0.98	1.09	± 0.0996		
96	96.01	0.00	0.01	± 0.0746		
100	98.99	-1.01	-1.01	± 0.0634		

#### **UUC\***: Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

FI-BME-NHS-CP-007/1 Rev.04 Page 3/3 Issued Date 20/07/2024