

Certificate No.

24PTP_0104240378

Page 1 of 2

Certificate of Calibration

EQUIPMENT : SPHYGMOMANOMETERS, ELECTRONIC, AUTOMATIC

ID CODE : PTP04634

MANUFACTURER : OMRON

MODEL : HEM-907

SERIAL No. : 20141000086AF

LOCATION : Check Up Center

SUBMITTED BY : PHYATHAI PHAHOLYOTHIN HOSPITAL

670/1 Phaholyothin Road, Samsen Nai, Phaya Thai, Bangkok 10400

Tel: 02-271-7000 Fax: -

CALIBRATED DATE : 24 APRIL 2024

ISSUE DATE : 1 MAY 2024

Kanittha.

Performed by : Approved by : KANITTHA SRIPRAPAD

21004

PHUWASIN YIWKIM

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATEVitalsign SimulatorFlukeProSim8254103123MD7777 Jun 2024

2. THIS CERTIFICATION IS TRACEABLE TO:

- Technology Promotion Association (Thai-Japan)

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Certificate No. 24PTP_0104240378

Page 2 of 2

Calibration Report

EQUIPMENT : SPHYGMOMANOMETERS, ELECTRONIC, AUTOMATIC

ID CODE : PTP04634

MANUFACTURER : OMRON

MODEL : HEM-907

SERIAL No. : 20141000086AF

DATE OF CALIBRATION : 24 APRIL 2024

ENVIRONMENT : TEMPERATURE 25 °C

RELATIVE HUMIDITY 55 %

PROCEDURE USED:

This instrument was calibration by comparison with standard

MEASURMENT RESULT:

/	Without Adjustment		Before Adjustment		After Adjustment
---	--------------------	--	-------------------	--	------------------

Systolic						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(mmHg)	(mmHg)	(mmHg)	(%)	(mmHg)		
60	60.00	0.00	0.00	± 0.9238		
120	120.00	0.00	0.00	± 1.2702		
200	198.00	-2.00	-1.00	± 1.7321		

Diastolic					
Standard Setting	UUC* Reading	Error	% Error	Uncertainty	
(mmHg)	(mmHg)	(mmHg)	(%)	(mmHg)	
30	27.00	-3.00	-10.00	± 0.7506	
80	79.00	-1.00	-1.25	± 1.0393	
150	152.00	2.00	1.33	± 1.4434	

Heart Rate						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(BPM)	(BPM)	(BPM)	(%)	(BPM)		
60	60.00	0.00	0.00	± 0.0150		
80	80.00	0.00	0.00	± 0.0460		
120	120.00	0.00	0.00	± 0.0460		

UUC*: Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%