Certificate of Calibration

EQUIPMENT	:	ANALYZERS, PHYSIOLOGIC, BODY COMPOSITION
ID CODE	:	PYT3_05967
MANUFACTURER	:	ACCUNIQ
MODEL	:	BC380
SERIAL No.	:	AKDL0039-20170914
LOCATION	:	PWA LIFE CENTER
SUBMITTED BY	:	PHYATHAI 3 HOSPITAL
		111 Phet Kasem Rd., Pak Khlong Phasi Charoen, Phasi Charoen Bangkok 10160
		Tel : (662) 467-1111 Fax : (622) 467-1111
CALIBRATED DATE	:	17 JANUARY 2025
ISSUE DATE	:	25 JANUARY 2025

Performed by : Peeranut _____ Approved by : _____

PEERANUT SIRISOMBOON *PHUWASIN YIWKIM*

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTER	MANUFACTURER	MODEL	SERIAL NO.	CERTIFICATE NO.	DUE DATE
Standard Weight 1 kg to 20 kg	LS	M3, 1kg - 120 kg	-	24-071060	16 Jun 2025

2. THIS CERTIFICATION IS TRACEABLE TO :

- ASIA MEDICAL AND AGRICULTURAL LABORATORY AND RESEARCH CENTER CO.,LTD

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY

Calibration Report

EQUIPMENT	:	ANALYZERS, PHYSIOLOGIC, BODY COMPOSITION	
ID CODE	:	PYT3_05967	
MANUFACTURER	:	ACCUNIQ	
MODEL	:	BC380	
SERIAL No.	:	AKDL0039-20170914	
DATE OF CALIBRATION	:	17 JANUARY 2025	
ENVIRONMENT	:	TEMPERATURE	25 °C
		RELATIVE HUMIDITY	50 %

PROCEDURE USED:

This instrument was calibration by comparison with standard

MEASURMENT RESULT:

[/] Without Adjustment [ ] Before Adjustment [ ] After Adjustment

Weight (Adult)				
Standard Setting (Kg)	UUC * Reading (Kg)	Error (Kg)	% Error (%)	Uncertainty (Kg)
20	19.93	-0.08	-0.38	± 0.0829
40	40.00	0.00	0.00	± 0.0577
60	59.98	-0.02	-0.04	± 0.0829
80	79.98	-0.03	-0.03	± 0.0829
100	99.95	-0.05	-0.05	± 0.0910
120	119.93	-0.07	-0.06	± 0.0829

Note: LOCATION : PWA LIFE CENTER (Examination room)

UUC* : Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

End of Report

2301/2 New Petchburi Soi 47 (soonvijai), Bangkapi, Huaykwang, Bangkok 10310
Tel: (66) 2762 4000 Fax: (66) 2762 4072