

Certificate No.

24PTP_0104240258

Page 1 of 2

Certificate of Calibration

EQUIPMENT : THERMOMETERS, ELECTRONIC, INFRARED, SKIN

ID CODE : PTP02981

MANUFACTURER : MICROLIFE

MODEL : FR1MF1

SERIAL No. : 04160881

LOCATION : Check Up Center

SUBMITTED BY : PHYATHAI PHAHOLYOTHIN HOSPITAL

670/1 Phaholyothin Road, Samsen Nai, Phaya Thai, Bangkok 10400

Tel: 02-271-7000 Fax: -

CALIBRATED DATE : 23 APRIL 2024

ISSUE DATE : 1 MAY 2024

Kanittha.

Performed by : KANITTHA SRIPRAPAD

3,110,24

PHUWASIN YIWKIM

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

Approved by:

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATEInfraredTEMPSENSCALsys37BBC/37BB/2021/552T24-00043 Jan 2025

Temperature Calibrator

2. THIS CERTIFICATION IS TRACEABLE TO:

- Unithai Group

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Certificate No. 24PTP_0104240258

Page 2 of 2

Calibration Report

EQUIPMENT : THERMOMETERS, ELECTRONIC, INFRARED, SKIN

ID CODE : PTP02981

MANUFACTURER : MICROLIFE

MODEL : FR1MF1
SERIAL No. : 04160881

DATE OF CALIBRATION : 23 APRIL 2024

ENVIRONMENT : TEMPERATURE 25 °C

RELATIVE HUMIDITY 55 %

PROCEDURE USED:

This instrument was calibration by comparison with standard

MEASURMENT RESULT:

/	Without Adjustment		Before Adjustment		After Adjustment
---	--------------------	--	-------------------	--	------------------

Temperature								
Standard Setting	UUC * Reading	Error	% Error	Uncertainty				
(°C)	(°C)	(°C)	(%)	(°C)				
35	35.33	0.32	0.93	± 0.1146				
37	37.15	0.15	0.41	± 0.3100				
39	39.18	0.18	0.45	± 0.1146				

UUC*: Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

2301/2 New Petchburi Soi 47 (soonvijai), Bangkapi, Huaykwang, Bangkok 10310 Tel: (66) 2762 4000 Fax: (66) 2762 4072

FI-BME-NHS-CP-061/1 Rev.03 Page 2/2 Issued Date 20/07/2023