

Certificate No. 24PTP_0104240364

Page 1 of 2

Certificate of Calibration

EQUIPMENT : WARMING/COOLING UNITS, PATIENT, CIRCULATING-

LIQUID

ID CODE : PTP02532

MANUFACTURER : CSZ

MODEL : BLANKETROL IIIjavascript:void(0);

SERIAL No. : 083-19368

LOCATION : I.C.U

SUBMITTED BY : PHYATHAI PHAHOLYOTHIN HOSPITAL

670/1 Phaholyothin Road, Samsen Nai, Phaya Thai, Bangkok 10400

Tel: 02-271-7000 Fax:-

CALIBRATED DATE : 26 APRIL 2024

ISSUE DATE : 1 MAY 2024

Kanittha.

KANITTHA SRIPRAPAD

Approved by:

PHUWASIN YIWKIM

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATETemperature TesterTESTO440 dP83962413DT2401035424 Jan 2025

2. THIS CERTIFICATION IS TRACEABLE TO:

- N-Health

Performed by:

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Certificate No. 24PTP_0104240364

Page 2 of 2

Calibration Report

EQUIPMENT : WARMING/COOLING UNITS, PATIENT, CIRCULATING-

LIQUID

ID CODE : PTP02532

MANUFACTURER : CSZ

MODEL : BLANKETROL III javascript:void(0);

SERIAL No. : 083-19368

DATE OF CALIBRATION : 26 APRIL 2024

ENVIRONMENT : TEMPERATURE 25 °C

RELATIVE HUMIDITY 55 %

PROCEDURE USED:

This instrument was calibration by comparison with standard

MEASURMENT RESULT:

/ Without Adjustment	Before Adjustment	After Adjustment
----------------------	-------------------	------------------

Fluid Temperature Display (Manual Mode)					
UUC* Setting	Standard Reading	Error	% Error	Uncertainty	
(°C)	(°C)	(°C)	(%)	(°C)	
10	10.33	0.32	3.25	± 0.3041	
30	30.13	0.13	0.42	± 0.3041	
37	37.05	0.05	0.14	± 0.3055	
40	39.88	-0.13	-0.31	± 0.4031	

UUC*: Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%

FI-BME-NHS-CP-066/1 Rev.03 Page 2/2 Issued Date 20/07/2023