

Report No. 25PYT3_01PM01250295

Page 1 of 2

Preventive Maintenance Action Report

EQUIPMENT : ANALYZERS, PHYSIOLOGIC, BODY COMPOSITION

ID CODE : PYT3_06536

MANUFACTURER : INBODY

MODEL : BSM370

SERIAL No. : 981900424

LOCATION : CHECK-UP

SUBMITTED BY : PHYATHAI 3 HOSPITAL

111 Phet Kasem Rd., Pak Khlong Phasi Charoen, Phasi Charoen Bangkok 10160

Tel: (662) 467-1111 Fax: (622) 467-1111

PM. DATE : 17 JANUARY 2025 ISSUE DATE : 25 JANUARY 2025

Supatcha

Performed by : _____ Approved by : _____

SUPATCHA BUAPUEAN PHUWASIN YIWKIM

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATEElectrical SafetyFlukeESA615349403424MD83710 Jun 2025

Analyzer

2. THIS CERTIFICATION IS TRACEABLE TO:

- Technology Promotion Association (Thai-Japan)

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Report No. 25PYT3_01PM01250295

Page 2 of 2

Preventive Maintenance Action Report

EQUIPMENT : ANALYZERS, PHYSIOLOGIC, BODY COMPOSITION

ID CODE : PYT3_06536

SERIAL No. : 981900424

1. Qualitative Tests

No.	List	Pass	Fail	Comment
1	Chassis/Housing/Frame	/		
2	Mount/Fasteners	/		
3	Casters/Wheels	/		
4	AC Plug	/		
5	Line Cord	/		
6	Strain Reliefs	/		
7	Controls/Switches	/		
8	Labeling	/		
9	Accessories	/		

2. Quantitative Tests

No.	List	Pass	Fail	Comment
1	Chassis Leakage Current ≤500 μA	/		14.2 μΑ
2	Zero Calibration/Electronic Scales Display reads	/		
	zero			
3	Weight (Mass) Accuracy for Electronic Scales ±1	/		
	%			

3. Preventive Maintenance

No.	List	Pass	Fail	Comment
1	Clean the exterior and accessories	/		

End of Report