

Certificate No.

24PYT3\_DEC12240021

Page 1 of 3

### **Certificate of Calibration**

EQUIPMENT : MONITORS, PATIENT TRANSPORT

ID CODE : PYT3\_02036

MANUFACTURER : PHILIPS

MODEL : SURESIGNS VSI

SERIAL No. : CN21207511

LOCATION : WARD 15

SUBMITTED BY : PHYATHAI 3 HOSPITAL

111 Phet Kasem Rd., Pak Khlong Phasi Charoen, Phasi Charoen Bangkok 10160

Tel: (662) 467-1111 Fax: (622) 467-1111

CALIBRATED DATE : 1 DECEMBER 2024

ISSUE DATE : 30 DECEMBER 2024

Performed by : Approved by : PUKARIN TONGKLIANG

SOM KONKAEW PUKARIN TONGKLIANG

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

### **CONDITION OF THIS RESULT OF TEST**

#### 1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATETester, Vital SignFLUKEPROSIM 8570205924MD54530 Apr 2025

Simulator

### 2. THIS CERTIFICATION IS TRACEABLE TO:

- Technology Promotion Association (Thai-Japan)

## 3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Certificate No. 24PYT3\_DEC12240021

Page 2 of 3

# **Calibration Report**

EQUIPMENT : MONITORS, PATIENT TRANSPORT

ID CODE : PYT3\_02036

MANUFACTURER : PHILIPS

MODEL : SURESIGNS VSI

SERIAL No. : CN21207511

DATE OF CALIBRATION : 1 DECEMBER 2024

ENVIRONMENT : TEMPERATURE 25 °C

RELATIVE HUMIDITY 55 %

### **PROCEDURE USED:**

This instrument was calibration by comparison with standard

### **MEASURMENT RESULT:**

/	Without Adjustment		Before Adjustment		After Adjustment
---	--------------------	--	-------------------	--	------------------

Systolic						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(mmHg)	(mmHg)	(mmHg)	(%)	(mmHg)		
60	60.02	0.02	0.03	± 0.0302		
120	116.02	-3.98	-3.31	± 0.0283		
200	193.25	-6.75	-3.38	± 0.8275		

Diastolic						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(mmHg)	(mmHg)	(mmHg)	(%)	(mmHg)		
30	30.97	0.97	3.24	± 0.0456		
80	79.00	-1.00	-1.25	± 0.0541		
150	144.98	-5.02	-3.35	± 0.0391		

Heart Rate						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(BPM)	(BPM)	(BPM)	(%)	(BPM)		
60	60.00	-0.00	-0.00	± 0.0600		
80	80.01	0.01	0.02	± 0.1131		
120	119.01	-0.99	-0.83	± 0.1163		

FI-BME-NHS-CP-007/1 Rev.04 Page 2/3 Issued Date 20/07/2024



Certificate No. 24PYT3\_DEC12240021

Page 3 of 3

# **Calibration Report**

EQUIPMENT : MONITORS, PATIENT TRANSPORT

ID CODE : PYT3\_02036

SERIAL No. : CN21207511

%Spo2						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(%SPO <sub>2</sub> )	(%SPO <sub>2</sub> )	(%SPO <sub>2</sub> )	(%)	(%SPO <sub>2</sub> )		
90	91.00	1.00	1.11	± 0.0746		
96	96.75	0.75	0.78	± 0.8275		
100	100.01	0.01	0.01	± 0.0557		

### **UUC\*: Unit Under Calibration**

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

FI-BME-NHS-CP-007/1 Rev.04 Page 3/3 Issued Date 20/07/2024