

Certificate No. 24PYT3_DEC12240111

Page 1 of 2

Certificate of Calibration

EQUIPMENT : ASPIRATORS, EMERGENCY

ID CODE : PYT3_01938

MANUFACTURER : CA-MI

MODEL : ASKIR 30

SERIAL No. : 44789

LOCATION : EENT(หูตาคอามูก)

SUBMITTED BY : PHYATHAI 3 HOSPITAL

111 Phet Kasem Rd., Pak Khlong Phasi Charoen, Phasi Charoen Bangkok 10160

Tel: (662) 467-1111 Fax: (622) 467-1111

CALIBRATED DATE : 3 DECEMBER 2024
ISSUE DATE : 30 DECEMBER 2024

In

Performed by:		Approved by :	1		
	SOM KONKAFW		PUKARIN TONGKI IANG		

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATETESTER,FLUKE717-30G59750045L202404137-000130 Apr 2025

PRESSURE CALIBRATOR

2. THIS CERTIFICATION IS TRACEABLE TO:

- Temperature, Dimension, mass, Anemometer, Electrical/Frequency, Relative Humidity

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Certificate No. 24PYT3_DEC12240111

Page 2 of 2

Calibration Report

EQUIPMENT : ASPIRATORS, EMERGENCY

ID CODE : PYT3_01938

MANUFACTURER : CA-MI

MODEL : ASKIR 30

SERIAL No. : 44789

DATE OF CALIBRATION : 3 DECEMBER 2024

ENVIRONMENT : TEMPERATURE 25 °C

RELATIVE HUMIDITY 55 %

PROCEDURE USED:

This instrument was calibration by comparison with standard

MEASURMENT RESULT:

/	Without Adjustment		Before Adjustment		After Adjustment
---	--------------------	--	-------------------	--	------------------

VACUUM								
UUC* Setting	Standard Reading	Error	% Error	Uncertainty				
(mmHg)	(mmHg)	(mmHg)	(%)	(mmHg)				
20	18.97	-1.03	-5.14	± 0.0625				
40	39.40	-0.60	-1.51	± 0.0435				
60	57.93	-2.07	-3.45	± 0.0135				
80	76.43	-3.58	-4.47	± 0.0627				

UUC*: Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

FI-BME-NHS-CP-015/1 Rev.04 Page 2/2 Issued Date 20/07/2024