Certificate of Calibration

EQUIPMENT	:	OXIMETERS, PULSE
ID CODE	:	PTP03287
MANUFACTURER	:	KTMED
MODEL	:	OXYPLAM
SERIAL No.	:	LK1401902
LOCATION	:	Diabetes Mellitus & Endocrine Center
SUBMITTED BY	:	PHYATHAI PHAHOLYOTHIN HOSPITAL
		670/1 Phaholyothin Road, Samsen Nai, Phaya Thai, Bangkok 10400
		Tel : 02-271-7000 Fax : -
CALIBRATED DATE	:	29 APRIL 2024
ISSUE DATE	:	1 MAY 2024

Performed by : _____ Approved by : _____

           SUPATCHA BUAPUEAN   *PHUWASIN YIWKIM*

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTER	MANUFACTURER	MODEL	SERIAL NO.	CERTIFICATE NO.	DUE DATE
Vitalsign Simulator	Fluke	ProSim8	2541031	23MD777	7 Jun 2024

2. THIS CERTIFICATION IS TRACEABLE TO :
   - Technology Promotion Association (Thai-Japan)

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY

Calibration Report

EQUIPMENT	:	OXIMETERS, PULSE
ID CODE	:	PTP03287
MANUFACTURER	:	KTMED
MODEL	:	OXYPLAM
SERIAL No.	:	LK1401902
DATE OF CALIBRATION	:	29 APRIL 2024
ENVIRONMENT	:	TEMPERATURE 24 °C
		RELATIVE HUMIDITY 55 %

PROCEDURE USED:

This instrument was calibration by comparison with standard

MEASURMENT RESULT:

☑ Without Adjustment   ☐ Before Adjustment   ☐ After Adjustment

%Spo2				
Standard Setting (%SPO$_2$)	UUC* Reading (%SPO$_2$)	Error (%SPO$_2$)	% Error (%)	Uncertainty (%SPO$_2$)
90	90.00	0.00	0.00	± 0.1039
96	96.00	0.00	0.00	± 0.1109
100	100.00	0.00	0.00	± 0.1155

Heart Rate				
Standard Setting (BPM)	UUC* Reading (BPM)	Error (BPM)	% Error (%)	Uncertainty (BPM)
60	60.00	0.00	0.00	± 0.0150
80	80.00	0.00	0.00	± 0.0460
120	120.00	0.00	0.00	± 0.0460

UUC* : Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%