

Report No. 24PYT3_I

24PYT3_DECPM12240015

Page 1 of 3

Preventive Maintenance Action Report

	SOM KONKAFW		PUKARIN TONGKLIANG
Performed by :	Gfm	Approved by:	Dung-
ISSUE DATE	: 30 DECE	MBER 2024	
PM. DATE	: 1 DECEM	MBER 2024	
	Tel: (662)) 467-1111 Fax : (62	2) 467-1111
	111 Phet Ka	sem Rd., Pak Khlong Phasi	Charoen, Phasi Charoen Bangkok 1016
SUBMITTED BY	: РНҮАТН	IAI 3 HOSPITAL	
LOCATION	: CHECK-U	UP (ตรวจสุขภาพ)	
SERIAL No.	: CN42755	215	
MODEL	: SURESIG	GNS VS4	
MANUFACTURER	: PHILIPS		
ID CODE	: PYT3_05	571	
EQUIPMENT	: MONITO	RS, PHYSIOLOGIC,	VITAL SIGNS

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTER MANUFACTURER MODEL SERIAL NO. CERTIFICATE NO. DUE DATE

2. THIS CERTIFICATION IS TRACEABLE TO:

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Report No. 24PYT3_DECPM12240015

Page 2 of 3

Preventive Maintenance Action Report

EQUIPMENT : MONITORS, PHYSIOLOGIC, VITAL SIGNS

ID CODE : PYT3_05571 SERIAL No. : CN42755215

1. Qualitative Tests

No.	List	Pass	Fail	None	Comment
1	Chassis/Housing	/			
2	Mount/Fasteners	/			
3	Casters/Brakes	/			
4	AC Plug	/			
5	Line Cord	/			
6	Strain Reliefs	/			
7	Circuit Breaker/Fuse	/			
8	Tubes/Bulbs	/			
9	Fittings/Connectors	/			
10	Transducers	/			
11	Controls/Switches	/			
12	Battery/Charger	/			
13	Time/Date Settings	/			
14	Cybersecurity			/	
15	User Calibration/Self-Test			/	
16	Alarms	/			
17	Labeling	/			
18	Accessories	/			

2. Quantitative Tests

No.	List	Pass	Fail	None	Comment
1	Grounding Resistance ≤0.5 Ω	/			0.380 Ω
2	Chassis Leakage Current ≤500 µA	/			126 μΑ
3	Air Leakage ≤15 mmHg/min	/			
4	Static Pressure Accuracy ±3 mmHg or	/			
	*Oscillometric Pressure ±8 mmHg				
5	Maximum Pressure (NIBP) ≤330 mmHg	/			
6	Heart Rate ±10%	/			

3. Preventive Maintenance

No.	List	Pass	Fail	None	Comment
1	Clean the exterior and accessories	/			
2	Lubricate			/	



Report No. 24PYT3_DECPM12240015

Page 3 of 3

Preventive Maintenance Action Report

EQUIPMENT : MONITORS, PHYSIOLOGIC, VITAL SIGNS

ID CODE : PYT3_05571 SERIAL No. : CN42755215

3. Preventive Maintenance (Continues)

No.	List	Pass	Fail	None	Comment
3	Adjust			/	
4	Replace			/	

FI-BME-NHS-CP-007/3 Rev.04 Page 3/3 Issued Date 20/07/2024