

Report No.

24PTP_01PM04240397

Page 1 of 2

Preventive Maintenance Action Report

EQUIPMENT : THERMOMETERS, ELECTRONIC, INFRARED, SKIN

ID CODE : PTP03305

MANUFACTURER : MICROLIFE

MODEL : FR1DL1

SERIAL No. :

LOCATION : Check Up Center

SUBMITTED BY : PHYATHAI PHAHOLYOTHIN HOSPITAL

670/1 Phaholyothin Road, Samsen Nai, Phaya Thai, Bangkok 10400

Tel: 02-271-7000 Fax:-

PM. DATE : 30 APRIL 2024

ISSUE DATE : 1 MAY 2024

Supatcha

Performed by : _____ Approved by : _____

SUPATCHA BUAPUEAN PHUWASIN YIWKIM

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTER MANUFACTURER MODEL SERIAL NO. CERTIFICATE NO. DUE DATE

2. THIS CERTIFICATION IS TRACEABLE TO:

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Report No. 24PTP_01PM04240397

Page 2 of 2

Preventive Maintenance Action Report

EQUIPMENT : THERMOMETERS, ELECTRONIC, INFRARED, SKIN

ID CODE : PTP03305

SERIAL No. : -

1. Qualitative Tests

No.	List	Pass	Fail	None	Comment
1	Chassis/Housing	/			
2	AC Plug			/	
3	Line Cord			/	
4	Strain Reliefs			/	
5	Cables			/	
6	Connectors			/	
7	Probes	/			
8	Controls/Switches	/			
9	Battery/Charger	/			
10	Indicators/Displays	/			
11	Self-Test Verification	/			
12	Time/Date Settings	/			
13	Cybersecurity			/	
14	Alarms	/			
15	Labeling	/			
16	Accessories	/			

2. Quantitative Tests

No.	List	Pass	Fail	None	Comment
1	Grounding Resistance ≤0.5 Ω			/	
2	Chassis Leakage Current ≤500 µA			/	
3	Temperature Accuracy ±0.5°C	/			
4	Temperature Alarms ±0.6°C	/			

3. Preventive Maintenance

	No.	List	Pass	Fail	None	Comment
	1	Clean the exterior and accessories	/			
	2	Lubricate			/	
	3	Adjust			/	
ſ	4	Replace			/	

Electrical Safety Test: Master used Electrical Safety Analyzer (Metron / QA-90) according to IEC.60601.1