

Certificate No.

24PTP_0104240062

Page 1 of 2

Certificate of Calibration

EQUIPMENT : INFUSION PUMPS

ID CODE : PTP02355

MANUFACTURER : B BRAUN

MODEL : INFUSOMAT SPACE P

SERIAL No. : 306228

LOCATION : Emergency & Ambulance

SUBMITTED BY : PHYATHAI PHAHOLYOTHIN HOSPITAL

670/1 Phaholyothin Road, Samsen Nai, Phaya Thai, Bangkok 10400

Tel: 02-271-7000 Fax: -

CALIBRATED DATE : 16 APRIL 2024

ISSUE DATE : 3 MAY 2024

Chakkaphan

CHAKKAPHAN KHAMPHUMI

Performed by:

Approved by :

PHUWASIN YIWKIM

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATEInfusion PumpRIGEL MedicalMulti-Flo23Q-141223MD116924 Aug 2024

Tester

2. THIS CERTIFICATION IS TRACEABLE TO:

- Technology Promotion Association (Thai-Japan)

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Certificate No. 24PTP_0104240062

Page 2 of 2

QUIPMENT	: INI	FUSION PUMPS			
O CODE	: PT	P02355			
IANUFACTURER	: B E	BRAUN			
IODEL	: INJ	FUSOMAT SPACE P			
ERIAL No.	: 306	5228			
ATE OF CALIBRA	ATION : 16	16 APRIL 2024			
NVIRONMENT		MPERATURE	24 °C		
TV IKOTVILIVI		LATIVE HUMIDITY	55 %		
/ Without Adjus	tment	Before Adjustment	After	Adjustmen	t
UUC* Setting	Standard Reading	Error	% Error	Unc	ertainty
(ml/hr)	(ml/hr)	(ml/hr)	(%)		nl/hr)
5	5.12	0.12	2.40	± ().0600
80	81.65	1.65	2.06	± (0.2600
200	205.74	5.74	2.87	± (0.7100
Occlusion Pressure	e < 1000 mmHg]			
UUC* Setting	Standard Reading				
	(mmHg)				
(mmHg)	745.00				
_				20	ml.
(mmHg)	B BRAUN	Drops/	/ml:	20	_