

Certificate No.

24PTP_0104240081

Page 1 of 2

Certificate of Calibration

EQUIPMENT : WARMING UNITS, PATIENT, RADIANT, INFANT

ID CODE : PTP01074

MANUFACTURER : FISHER & PAYKEL HEALTHCARE

MODEL : COSYCOT

SERIAL No. : 151208000753

LOCATION : Nursery

SUBMITTED BY : PHYATHAI PHAHOLYOTHIN HOSPITAL

670/1 Phaholyothin Road, Samsen Nai, Phaya Thai, Bangkok 10400

Tel: 02-271-7000 Fax: -

CALIBRATED DATE : 17 APRIL 2024

ISSUE DATE : 3 MAY 2024

Supatcha		むりゃく
	Approved by:	

SUPATCHA BUAPUEAN PHUWASIN YIWKIM

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATEINCUBATORFLUKEINCU II54850013L20230525126 May 2024

AND RADIANT BIOMEDICAL

WARMER

Performed by:

ANALYZER

Pressure Calibrator Fluke 717 30G 5730135 23P1956 12 Jun 2024

2. THIS CERTIFICATION IS TRACEABLE TO:

- $Temperature, Dimension, mass, Anemometer, Electrical/Frequency, Relative\ Humidity$
- Technology Promotion Association (Thai-Japan)

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Certificate No. 24PTP_0104240081

Page 2 of 2

Calibration Report

EQUIPMENT : WARMING UNITS, PATIENT, RADIANT, INFANT

ID CODE : PTP01074

MANUFACTURER : FISHER & PAYKEL HEALTHCARE

MODEL : COSYCOT

SERIAL No. : 151208000753

DATE OF CALIBRATION : 17 APRIL 2024

ENVIRONMENT : TEMPERATURE 24 °C

RELATIVE HUMIDITY 56 %

PROCEDURE USED:

This instrument was calibration by comparison with standard

MEASURMENT RESULT:

l	/	Without Adjustment	Before Adjustment	After Adjustment

Skin Temperature Control					
UUC* Setting	Standard Reading	Error	% Error	Uncertainty	
(°C)	(°C)	(°C)	(%)	(°C)	
36	35.90	-0.10	-0.28	± 0.0793	

Pressure					
UUC* Setting	Standard Reading	Error	% Error	Uncertainty	
(cmH ₂ O)	(cmH ₂ O)	(cmH ₂ O)	(%)	(cmH ₂ O)	
10	9.65	-0.35	-3.50	± 0.5129	
20	20.25	0.25	1.25	± 0.5063	
30	30.88	0.88	2.92	± 0.5121	
40	39.85	-0.15	-0.37	± 0.5129	

UUC*: Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%

FI-BME-NHS-CP-028/1 Rev.03 Page 2/2 Issued Date 20/07/2023