

Report No.

24PYT3\_DECPM12240165

Page 1 of 2

# **Preventive Maintenance Action Report**

	SOM VONVAEW	DUVADIN TONCVI IANC
Performed by:	Gfr	Approved by :
ISSUE DATE	: 30 DECEMI	BER 2024
PM. DATE	: 16 DECEMI	BER 2024
	Tel: (662) 4	167-1111 Fax: (622) 467-1111
	111 Phet Kasen	n Rd., Pak Khlong Phasi Charoen, Phasi Charoen Bangkok 1016
SUBMITTED BY	: PHYATHAI	I 3 HOSPITAL
LOCATION	: NURSERY(	(เด็กอ่อน)
SERIAL No.	: 178109	
MODEL	: RADICAL-7	7
MANUFACTURE	: MASIMO	
ID CODE	: PYT3_0161	7
EQUIPMENT	: OXIMETER	RS, PULSE

#### **CONDITION OF THIS RESULT OF TEST**

1. REFERENCE STANDARD INSTRUMENT:

MASTER MANUFACTURER MODEL SERIAL NO. CERTIFICATE NO. DUE DATE

2. THIS CERTIFICATION IS TRACEABLE TO:

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



*Report No.* 24PYT3\_DECPM12240165

Page 2 of 2

## **Preventive Maintenance Action Report**

EQUIPMENT : OXIMETERS, PULSE

ID CODE : PYT3\_01617

SERIAL No. : 178109

### 1. Qualitative Tests

No.	List	Pass	Fail	None	Comment
1	Chassis/Housing	/			
2	2 Mount/Fasteners			/	
3	AC Plug/Receptacles			/	
4	Line Cord			/	
5	Strain Reliefs			/	
6	Circuit Breaker/Fuse			/	
7	Cables			/	
8	Connectors	/			
9	Probes/Sensors/Transducers	/			
10	Controls/Switches	/			
11	Battery/Charger	/			
12	Indicators/Displays	/			
13	Self-Test	/			
14	Time/Date Settings	/			
15	Cybersecurity			/	
16	Alarms	/			
17	Audible Signals	/			
18	Labeling	/			

#### 2. Quantitative Tests

No.	List	Pass	Fail	None	Comment
1	Grounding Resistance ≤0.5 Ω	/			0.251 Ω
2	Chassis Leakage Current ≤500 µA	/			190.3 μΑ

### 3. Preventive Maintenance

No.	List	Pass	Fail	None	Comment
1	Clean the exterior and accessories	/			
2	Lubricate			/	
3	Adjust			/	
4	Replace			/	