

Report No. 24PYT3\_DECPM12240001

Page 1 of 2

# **Preventive Maintenance Action Report**

**EQUIPMENT** : OXIMETERS, PULSE

ID CODE : PYT3\_05003

MANUFACTURER : EDAN

**MODEL** : H100B

SERIAL No. : 113575-M16906130010

LOCATION : HOPD CENTER

SUBMITTED BY : PHYATHAI 3 HOSPITAL

111 Phet Kasem Rd., Pak Khlong Phasi Charoen, Phasi Charoen Bangkok 10160

Tel: (662) 467-1111 Fax: (622) 467-1111

PM. DATE : 1 DECEMBER 2024

**ISSUE DATE** : 30 DECEMBER 2024

Performed by:

PUKARIN TONGKLIANG

#### **CONDITION OF THIS RESULT OF TEST**

1. REFERENCE STANDARD INSTRUMENT:

MASTER MANUFACTURER MODEL SERIAL NO. CERTIFICATE NO. DUE DATE

2. THIS CERTIFICATION IS TRACEABLE TO:

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF **TEST ONLY** 



Report No. 24PYT3\_DECPM12240001

Page 2 of 2

## **Preventive Maintenance Action Report**

EQUIPMENT : OXIMETERS, PULSE

ID CODE : PYT3\_05003

SERIAL No. : 113575-M16906130010

### 1. Qualitative Tests

| No. | List                       | Pass | Fail | None | Comment |
|-----|----------------------------|------|------|------|---------|
| 1   | Chassis/Housing            | /    |      |      |         |
| 2   | Mount/Fasteners            |      |      | /    |         |
| 3   | AC Plug/Receptacles        |      |      | /    |         |
| 4   | Line Cord                  |      |      | /    |         |
| 5   | Strain Reliefs             |      |      | /    |         |
| 6   | Circuit Breaker/Fuse       |      |      | /    |         |
| 7   | Cables                     |      |      | /    |         |
| 8   | Connectors                 | /    |      |      |         |
| 9   | Probes/Sensors/Transducers | /    |      |      |         |
| 10  | Controls/Switches          | /    |      |      |         |
| 11  | Battery/Charger            | /    |      |      |         |
| 12  | Indicators/Displays        | /    |      |      |         |
| 13  | Self-Test                  | /    |      |      |         |
| 14  | Time/Date Settings         | /    |      |      |         |
| 15  | Cybersecurity              |      |      | /    |         |
| 16  | Alarms                     | /    |      |      |         |
| 17  | Audible Signals            | /    |      |      |         |
| 18  | Labeling                   | /    |      |      |         |

#### 2. Quantitative Tests

| No. | List                            | Pass | Fail | None | Comment |
|-----|---------------------------------|------|------|------|---------|
| 1   | Grounding Resistance ≤0.5 Ω     |      |      | /    |         |
| 2   | Chassis Leakage Current ≤500 µA |      |      | /    |         |

### 3. Preventive Maintenance

| No. | List                               | Pass | Fail | None | Comment |
|-----|------------------------------------|------|------|------|---------|
| 1   | Clean the exterior and accessories | /    |      |      |         |
| 2   | Lubricate                          |      |      | /    |         |
| 3   | Adjust                             |      |      | /    |         |
| 4   | Replace                            |      |      | /    |         |