

Report No.

24PTP\_01PM04240130

Page 1 of 2

## **Preventive Maintenance Action Report**

EQUIPMENT : MODULE, ECG/SPO2/NIBP/IBP/TEMP

ID CODE : PTP02719

MANUFACTURER : PHILIPS

MODEL : M3012A

SERIAL No. : DE83753516

LOCATION : I.C.U

SUBMITTED BY : PHYATHAI PHAHOLYOTHIN HOSPITAL

670/1 Phaholyothin Road, Samsen Nai, Phaya Thai, Bangkok 10400

Tel: 02-271-7000 Fax:-

PM. DATE : 18 APRIL 2024

ISSUE DATE : 3 MAY 2024

Performed by:

Approved by:

21012 A

USANEE CHALONGPHOKSILCHAI PHUWASIN YIWKIM

## **CONDITION OF THIS RESULT OF TEST**

1. REFERENCE STANDARD INSTRUMENT:

MASTER MANUFACTURER MODEL SERIAL NO. CERTIFICATE NO. DUE DATE

2. THIS CERTIFICATION IS TRACEABLE TO:

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Report No. 24PTP\_01PM04240130

Page 2 of 2

## **Preventive Maintenance Action Report**

EQUIPMENT : MODULE, ECG/SPO2/NIBP/IBP/TEMP

ID CODE : PTP02719

SERIAL No. : DE83753516

## 1.Report

No.	List	Pass	Fail	None	Comment
1	This Module Calibrate With Monitor	/			

**2.Remark** MODULE OF ID Code :PTP00403

Electrical Safety Test: Master used Electrical Safety Analyzer (Metron / QA-90) according to IEC.60601.1

FI-BME-NHS-CP-001/3 Rev.09 Page 2/2 Issued Date 20/07/2023