

Certificate No.

24PYT3_DEC12240006

Page 1 of 3

Certificate of Calibration

EQUIPMENT : MONITORS, PATIENT TRANSPORT

ID CODE : PYT3_03000

MANUFACTURER : PHILIPS

MODEL : SURESIGNS VSI

SERIAL No. : CN21208240

LOCATION : WARD 17

SUBMITTED BY : PHYATHAI 3 HOSPITAL

111 Phet Kasem Rd., Pak Khlong Phasi Charoen, Phasi Charoen Bangkok 10160

Tel: (662) 467-1111 Fax: (622) 467-1111

CALIBRATED DATE : 1 DECEMBER 2024

ISSUE DATE : 30 DECEMBER 2024

Performed by : Approved by : PUKARIN TONGKLIANG

SOM KONKAEW PUKARIN TONGKLIANG

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATETester, Vital SignFLUKEPROSIM 8570205924MD54530 Apr 2025

Simulator

2. THIS CERTIFICATION IS TRACEABLE TO:

- Technology Promotion Association (Thai-Japan)

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Certificate No. 24PYT3_DEC12240006

Page 2 of 3

Calibration Report

EQUIPMENT : MONITORS, PATIENT TRANSPORT

ID CODE : PYT3_03000

MANUFACTURER : PHILIPS

MODEL : SURESIGNS VSI

SERIAL No. : CN21208240

DATE OF CALIBRATION : 1 DECEMBER 2024

ENVIRONMENT : TEMPERATURE 25 °C

RELATIVE HUMIDITY 55 %

PROCEDURE USED:

This instrument was calibration by comparison with standard

MEASURMENT RESULT:

/	Without Adjustment		Before Adjustment		After Adjustment
---	--------------------	--	-------------------	--	------------------

Systolic						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(mmHg)	(mmHg)	(mmHg)	(%)	(mmHg)		
60	61.75	1.75	2.92	± 0.8275		
120	117.25	-2.75	-2.29	± 0.8275		
200	194.25	-5.75	-2.88	± 0.8275		

Diastolic						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(mmHg)	(mmHg)	(mmHg)	(%)	(mmHg)		
30	32.75	2.75	9.17	± 0.8275		
80	80.25	0.25	0.31	± 0.8275		
150	144.25	-5.75	-3.83	± 0.8275		

Heart Rate						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(BPM)	(BPM)	(BPM)	(%)	(BPM)		
60	60.25	0.25	0.42	± 0.8280		
80	80.25	0.25	0.31	± 0.8473		
120	120.25	0.25	0.21	± 0.8473		

FI-BME-NHS-CP-007/1 Rev.04 Page 2/3 Issued Date 20/07/2024



Certificate No. 24PYT3_DEC12240006

Page 3 of 3

Calibration Report

EQUIPMENT : MONITORS, PATIENT TRANSPORT

ID CODE : PYT3_03000 SERIAL No. : CN21208240

%Spo2						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(%SPO ₂)	(%SPO ₂)	(%SPO ₂)	(%)	(%SPO ₂)		
90	90.75	0.75	0.83	± 0.8275		
96	96.75	0.75	0.78	± 0.8275		
100	99.75	-0.25	-0.25	± 0.8275		

UUC*: Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

FI-BME-NHS-CP-007/1 Rev.04 Page 3/3 Issued Date 20/07/2024