

Certificate No. 24PTP\_0104240114

Page 1 of 3

### **Certificate of Calibration**

EQUIPMENT : VENTILATORS, INTENSIVE CARE

ID CODE : PTP00258

MANUFACTURER : CAREFUSION

MODEL : VELA

SERIAL No. : BFT01867

LOCATION : I.C.U

SUBMITTED BY : PHYATHAI PHAHOLYOTHIN HOSPITAL

670/1 Phaholyothin Road, Samsen Nai, Phaya Thai, Bangkok 10400

Tel: 02-271-7000 Fax: -

CALIBRATED DATE : 18 APRIL 2024

ISSUE DATE : 3 MAY 2024

Supatcha

Performed by : Approved by : PHUWASIN YIWKIM

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

### **CONDITION OF THIS RESULT OF TEST**

#### 1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATEGas Flow AnalyzerFluke BiomedicalVT650606972523MD115827 Aug 2024

#### 2. THIS CERTIFICATION IS TRACEABLE TO:

- Technology Promotion Association (Thai-Japan)

# 3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Certificate No. 24PTP\_0104240114

Page 2 of 3

# **Calibration Report**

EQUIPMENT : VENTILATORS, INTENSIVE CARE

ID CODE : PTP00258

MANUFACTURER : CAREFUSION

MODEL : VELA

SERIAL No. : BFT01867

DATE OF CALIBRATION : 18 APRIL 2024

ENVIRONMENT : TEMPERATURE 24 °C

RELATIVE HUMIDITY 56 %

### PROCEDURE USED:

This instrument was calibration by comparison with standard

### **MEASURMENT RESULT:**

/	Without Adjustment		Before Adjustment		After Adjustment
---	--------------------	--	-------------------	--	------------------

TIDAL VOLUME						
UUC* Setting	Standard Reading	Error	% Error	Uncertainty		
(mL)	(mL)	(mL)	(%)	(mL)		
400	382.75	-17.25	-4.31	± 14.4454		
600	603.30	3.30	0.55	± 21.7457		
800	794.88	-5.13	-0.64	± 28.9400		

BREATH RATE						
UUC* Setting	Standard Reading	Error	% Error	Uncertainty		
(BPM)	(BPM)	(BPM)	(%)	(BPM)		
30	30.00	0.00	0.00	± 0.3464		
20	20.00	0.00	0.00	± 0.2309		
15	15.00	0.00	0.00	± 0.1732		

I:E RATIO						
UUC* Setting	Standard Reading	Error	% Error	Uncertainty		
((1:x))	((1:x))	((1:x))	(%)	((1:x))		
1	1.00	0.00	0.00	± 0.0231		
2	1.99	-0.01	-0.50	± 0.0462		
3	2.93	-0.07	-2.33	± 0.0693		

FI-BME-NHS-CP-018/1 Rev.03 Page 2/3 Issued Date 20/07/2023



*Certificate No.* 24PTP\_0104240114

Page 3 of 3

# **Calibration Report**

EQUIPMENT : VENTILATORS, INTENSIVE CARE

ID CODE : PTP00258 SERIAL No. : BFT01867

PIP						
UUC* Setting	Standard Reading	Error	% Error	Uncertainty		
(cmH <sub>2</sub> O)	(cmH <sub>2</sub> O)	(cmH <sub>2</sub> O)	(%)	(cmH <sub>2</sub> O)		
20	19.64	-0.36	-1.80	± 0.3200		
30	29.09	-0.91	-3.03	± 0.3200		
40	39.18	-0.82	-2.05	± 0.3200		

PEEP						
UUC* Setting	Standard Reading	Error	% Error	Uncertainty		
(cmH2O)	(cmH2O)	(cmH2O)	(%)	(cmH2O)		
5	4.70	-0.30	-6.00	± 0.0577		
10	10.30	0.30	3.00	± 0.1155		
15	14.89	-0.11	-0.73	± 0.1732		

% OXYGEN						
UUC* Setting	Standard Reading	Error	% Error	Uncertainty		
(%O <sub>2</sub> )	(%O <sub>2</sub> )	$(\%O_2)$	(%)	(%O <sub>2</sub> )		
21	21.70	0.70	3.33	± 0.4850		
60	60.30	0.30	0.50	± 1.3857		
100	99.20	-0.80	-0.80	± 2.3095		

### **UUC\***: Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

FI-BME-NHS-CP-018/1 Rev.03 Page 3/3 Issued Date 20/07/2023