

Report No.

25PYT3_01PM01250210

Page 1 of 2

Preventive Maintenance Action Report

EQUIPMENT : ANALYZERS, PHYSIOLOGIC, BODY COMPOSITION

ID CODE : PYT3_09071

MANUFACTURER : OMRON

MODEL : HBF-222T

SERIAL No. : 20220104872F

LOCATION : PHYSICAL THERAPY

SUBMITTED BY : PHYATHAI 3 HOSPITAL

111 Phet Kasem Rd., Pak Khlong Phasi Charoen, Phasi Charoen Bangkok 10160

Tel: (662) 467-1111 Fax: (622) 467-1111

PM. DATE : 16 JANUARY 2025 ISSUE DATE : 25 JANUARY 2025

Chakkaphan

CHAKKAPHAN KHAMPHUMI

Approved by :

PHUWASIN YIWKIM

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTER MANUFACTURER MODEL SERIAL NO. CERTIFICATE NO. DUE DATE

2. THIS CERTIFICATION IS TRACEABLE TO:

-

Performed by:

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Report No. 25PYT3_01PM01250210

Page 2 of 2

Preventive Maintenance Action Report

EQUIPMENT : ANALYZERS, PHYSIOLOGIC, BODY COMPOSITION

ID CODE : PYT3_09071

SERIAL No. : 20220104872F

1. Qualitative Tests

No.	List	Pass	Fail	Comment
1	Chassis/Housing/Frame	/		
2	Controls/Switches	/		
3	Battery/Charger	/		
4	Labeling	/		
5	Accessories	/		

2. Quantitative Tests

No.	List	Pass	Fail	Comment
1	Zero Calibration/Electronic Scales Display reads	/		
	zero			
2	Weight (Mass) Accuracy for Electronic Scales ±1	/		
	%			

3. Preventive Maintenance

No	. List	Pass	Fail	Comment
1	Clean the exterior and accessories	/		

End of Report