

Certificate No.

24PTP\_0104240100

Page 1 of 2

## **Certificate of Calibration**

EQUIPMENT : OXIMETERS, PULSE

ID CODE : PTP00492

MANUFACTURER : MASIMO

MODEL : RADICAL-7

SERIAL No. : 241129

LOCATION : Nursery

SUBMITTED BY : PHYATHAI PHAHOLYOTHIN HOSPITAL

670/1 Phaholyothin Road, Samsen Nai, Phaya Thai, Bangkok 10400

Tel: 02-271-7000 Fax: -

CALIBRATED DATE : 17 APRIL 2024

ISSUE DATE : 1 MAY 2024

Supatcha

Performed by : Approved by : PHUWASIN YIWKIM

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

### **CONDITION OF THIS RESULT OF TEST**

### 1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATEVitalsign SimulatorFlukeProSim8254103123MD7777 Jun 2024

#### 2. THIS CERTIFICATION IS TRACEABLE TO:

- Technology Promotion Association (Thai-Japan)

# 3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



*Certificate No.* 24PTP\_0104240100

Page 2 of 2

# **Calibration Report**

EQUIPMENT : OXIMETERS, PULSE

ID CODE : PTP00492

MANUFACTURER : MASIMO

MODEL : RADICAL-7

SERIAL No. : 241129

DATE OF CALIBRATION : 17 APRIL 2024

ENVIRONMENT : TEMPERATURE 24 °C

RELATIVE HUMIDITY 55 %

#### **PROCEDURE USED:**

This instrument was calibration by comparison with standard

### **MEASURMENT RESULT:**

/	Without Adjustment		Before Adjustment		After Adjustment
---	--------------------	--	-------------------	--	------------------

%Spo2									
Standard Setting	UUC* Reading	Error	% Error	Uncertainty					
(%SPO <sub>2</sub> )	(%SPO <sub>2</sub> )	(%SPO <sub>2</sub> )	(%)	(%SPO <sub>2</sub> )					
90	90.50	0.50	0.56	± 0.9709					
96	96.00	0.00	0.00	± 0.1109					
100	99.50	-0.50	-0.50	± 0.9744					

Heart Rate								
Standard Setting	UUC* Reading	Error	% Error	Uncertainty				
(BPM)	(BPM)	(BPM)	(%)	(BPM)				
60	60.00	0.00	0.00	± 0.0150				
80	80.00	0.00	0.00	± 0.0460				
120	120.00	0.00	0.00	± 0.0460				

### **UUC\*: Unit Under Calibration**

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%

FI-BME-NHS-CP-012/1 Rev.03 Page 2/2 Issued Date 20/07/2023