

Certificate No. 24

24PYT3\_DEC12240132

Page 1 of 2

### **Certificate of Calibration**

EQUIPMENT : ASPIRATORS, EMERGENCY

ID CODE : PYT3\_00595

MANUFACTURER : FOCUS

MODEL : VAC-420

SERIAL No. : -

LOCATION : AMBULANCE (รถพยาบาล)

SUBMITTED BY : PHYATHAI 3 HOSPITAL

111 Phet Kasem Rd., Pak Khlong Phasi Charoen, Phasi Charoen Bangkok 10160

Tel: (662) 467-1111 Fax: (622) 467-1111

CALIBRATED DATE : 4 DECEMBER 2024

ISSUE DATE : 30 DECEMBER 2024

Performed by : Approved by : PUKARIN TONGKLIANG

SOM KONKAEW PUKARIN TONGKLIANG

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

### **CONDITION OF THIS RESULT OF TEST**

### 1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATETESTER,FLUKE717-30G59750045L202404137-000130 Apr 2025

PRESSURE CALIBRATOR

#### 2. THIS CERTIFICATION IS TRACEABLE TO:

- Temperature, Dimension, mass, Anemometer, Electrical/Frequency, Relative Humidity

## 3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Certificate No. 24PYT3\_DEC12240132

Page 2 of 2

# **Calibration Report**

EQUIPMENT : ASPIRATORS, EMERGENCY

ID CODE : PYT3\_00595

MANUFACTURER : FOCUS

MODEL : VAC-420

SERIAL No. : -

DATE OF CALIBRATION : 4 DECEMBER 2024

ENVIRONMENT : TEMPERATURE 25 °C

RELATIVE HUMIDITY 55 %

#### **PROCEDURE USED:**

This instrument was calibration by comparison with standard

### **MEASURMENT RESULT:**

/	Without Adjustment		Before Adjustment		After Adjustment
---	--------------------	--	-------------------	--	------------------

VACUUM								
UUC* Setting	Standard Reading	Error	% Error	Uncertainty				
(mmHg)	(mmHg)	(mmHg)	(%)	(mmHg)				
200	208.16	8.16	4.08	± 0.0549				
300	304.21	4.21	1.40	± 0.0826				
400	402.76	2.76	0.69	± 0.0487				
500	500.85	0.85	0.17	± 0.0589				
600	597.18	-2.83	-0.47	± 0.0497				

### **UUC\***: Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

2301/2 New Petchburi Soi 47 (soonvijai), Bangkapi, Huaykwang, Bangkok 10310 Tel: (66) 2762 4000 Fax: (66) 2762 4072

FI-BME-NHS-CP-015/1 Rev.04 Page 2/2 Issued Date 20/07/2024