

Certificate No.

24PYT3\_DEC12240033

Page 1 of 2

### **Certificate of Calibration**

EQUIPMENT : OXIMETERS, PULSE

ID CODE : PYT3\_08674

MANUFACTURER : EDAN

MODEL : H100B

SERIAL No. : 360101-M22507370021

LOCATION : WARD 9

SUBMITTED BY : PHYATHAI 3 HOSPITAL

111 Phet Kasem Rd., Pak Khlong Phasi Charoen, Phasi Charoen Bangkok 10160

Tel: (662) 467-1111 Fax: (622) 467-1111

CALIBRATED DATE : 1 DECEMBER 2024

ISSUE DATE : 30 DECEMBER 2024

Performed by : Approved by : PUKARIN TONGKLIANG

SOM KONKAEW PUKARIN TONGKLIANG

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

### **CONDITION OF THIS RESULT OF TEST**

#### 1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATETester, Vital SignFLUKEPROSIM 8570205924MD54530 Apr 2025Simulator

### 2. THIS CERTIFICATION IS TRACEABLE TO:

- Technology Promotion Association (Thai-Japan)

# 3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Certificate No. 24PYT3\_DEC12240033

Page 2 of 2

# **Calibration Report**

EQUIPMENT : OXIMETERS, PULSE

ID CODE : PYT3\_08674

MANUFACTURER : EDAN MODEL : H100B

SERIAL No. : 360101-M22507370021

DATE OF CALIBRATION : 1 DECEMBER 2024

ENVIRONMENT : TEMPERATURE 25 °C

RELATIVE HUMIDITY 55 %

#### **PROCEDURE USED:**

This instrument was calibration by comparison with standard

### **MEASURMENT RESULT:**

/	Without Adjustment		Before Adjustment		After Adjustment
---	--------------------	--	-------------------	--	------------------

%Spo2								
Standard Setting	UUC* Reading	Error	% Error	Uncertainty				
(%SPO <sub>2</sub> )	(%SPO <sub>2</sub> )	(%SPO <sub>2</sub> )	(%)	(%SPO <sub>2</sub> )				
90	91.00	0.99	1.11	± 0.0416				
96	96.98	0.98	1.02	± 0.0547				
100	99.98	-0.02	-0.02	± 0.0769				

Heart Rate									
Standard Setting	UUC* Reading	Error	% Error	Uncertainty					
(BPM)	(BPM)	(BPM)	(%)	(BPM)					
60	60.01	0.01	0.02	± 0.0938					
80	80.01	0.01	0.01	± 0.1139					
120	119.97	-0.03	-0.03	± 0.1156					

### **UUC\*: Unit Under Calibration**

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%

FI-BME-NHS-CP-012/1 Rev.04 Page 2/2 Issued Date 20/07/2024