

Certificate No. 24PT

24PTP_0104240072

Page 1 of 2

Certificate of Calibration

EQUIPMENT : OXIMETERS, PULSE

ID CODE : PTP03803 MANUFACTURER : JUMPER

MODEL : JPD-500E

SERIAL No. : 8823500E408006

LOCATION : Patient Wards 406

SUBMITTED BY : PHYATHAI PHAHOLYOTHIN HOSPITAL

670/1 Phaholyothin Road, Samsen Nai, Phaya Thai, Bangkok 10400

Tel: 02-271-7000 Fax: -

CALIBRATED DATE : 17 APRIL 2024

ISSUE DATE : 1 MAY 2024

Performed by:

Approved by :

USANEE CHALONGPHOKSILCHAI

PHUWASIN YIWKIM

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATEVitalsign SimulatorFlukeProSim8254103123MD7777 Jun 2024

2. THIS CERTIFICATION IS TRACEABLE TO:

- Technology Promotion Association (Thai-Japan)

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Certificate No. 24PTP_0104240072

Page 2 of 2

Calibration Report

EQUIPMENT : OXIMETERS, PULSE

ID CODE : PTP03803

MANUFACTURER : JUMPER

MODEL : JPD-500E

SERIAL No. : 8823500E408006

DATE OF CALIBRATION : 17 APRIL 2024

ENVIRONMENT : TEMPERATURE 25 °C

RELATIVE HUMIDITY 55 %

PROCEDURE USED:

This instrument was calibration by comparison with standard

MEASURMENT RESULT:

/	Without Adjustment		Before Adjustment		After Adjustment
---	--------------------	--	-------------------	--	------------------

%Spo2								
Standard Setting	UUC* Reading	Error	% Error	Uncertainty				
(%SPO ₂)	(%SPO ₂)	(%SPO ₂)	(%)	(%SPO ₂)				
90	91.00	1.00	1.11	± 0.1039				
96	97.00	1.00	1.04	± 0.1109				
100	100.00	0.00	0.00	± 0.1155				

Heart Rate								
Standard Setting	UUC* Reading	Error	% Error	Uncertainty				
(BPM)	(BPM)	(BPM)	(%)	(BPM)				
60	59.00	-1.00	-1.67	± 0.0150				
80	79.00	-1.00	-1.25	± 0.0460				
120	119.00	-1.00	-0.83	± 0.0460				

UUC*: Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%

FI-BME-NHS-CP-012/1 Rev.03 Page 2/2 Issued Date 20/07/2023